ADVANCED TECHNOLOGICAL AIDS FOR THE HANDICAPPED

by

Leslie Howe Marley

B.S., Virginia Polytechnic Institute and State University, 1979

A MASTER'S REPORT

submitted in partial fulfillment of the

requirements for the degree

MASTER OF SCIENCE

Department of Computing and Information Sciences

KANSAS STATE UNIVERSITY Manhattan, Kansas

1988

Major Professor

CONTENTS Alleg 312035

CHAPTER 1	INTRODUCTION										1			
	I.1 Evaluating The Need										1			
	1.2 Report Organization										4			
CHAPTER 2 COMPUTER AIDS FOR THE HANDICAPPED														
	2.1 Introduction	Ċ	Ċ	i		i	i	Ċ	Ċ		5			
	2.2 Reduction of Handicap			i	i	i	ċ	÷	÷		5			
	2.2.1 Vision Impairment			÷	÷	÷	÷				6			
	2.2.1.1 Mobility Aids.										6			
											10			
	2.2.1.2.1 Reading Machines										10			
	2.2.1.2.2 Transient Braille										11			
											12			
	2.2.1.2.4 Braille Trainer										13			
	2.2.2 Hearing Impairment										14			
	2.2.2.1 Telecommunications										14			
	2.2.2.2 Programmable Hearing Aids										16			
	2.2.2.3 Tactile Stimulation										17			
	2.2.2.4 Teaching Aids										18			
	2.2.3 Motor Impairment										19			
	2.2.3.1 Communications										19			
	2.2.3.2 Environmental Control Systems										21			
	2.2.3.3 Robotic Aids										22			
	2.2.3.4 Entertainment										25			
	2.2.3.5 Summary										26			
	2.3 Replacement of Function										27			
	2.3.1 Visual Prosthesis										28			
	2.3.2 Cochlear Prosthesis										31			
	2.3.3 Functional Neuromuscular Stimulation.										34			
	2.3.4 Computer Utilization										38			
	2.3.5 Conclusion										38			
CHAPTER 3	WHEELCHAIRS										39			
01811 1811 0	3.1 Introduction							Ċ			39			
	3.2 History of The Wheelchair							Ċ			39			
	3.3 Present-Day Wheelchairs			Ċ		Ċ		Ċ	0		40			
	3.3.1 Nonfolding Chairs			•	•	•	•	•	•	•	40			
	3.3.2 Rearwheel-Drive Chairs						٠	•	•		41			
	3.3.3 Sports Chairs										41			
	3.3.4 Frontwheel-Drive Chairs			Ċ	Ċ		÷	Ċ			42			
	3.3.5 Amputee Chairs				Ċ	Ċ	Ċ				43			
	3.3.6 One-Arm-Drive Chairs.	•	•	•	•		•	•	•	•	43			
	3.3.7 Motorized Chairs			•	•		•			•	44			
	3.3.8 Costs				Ċ	1	Ċ				45			
	3.4 Lack of Innovation.				Ĺ	ĵ.	Ĺ	Ĺ			45			
	3.4.1 Consumer Market				í	í	í		í		45			
	3.4.2 Third-Party Reimbursers								í		46			
	3.4.3 Standards							í	í		47			

m -7

		3.4.4	Consur	ner U	nav	are	nes	5.															48
		3.4.5	Produc	t Lial	bilit	y.																	48
CHAPTER 4	DE	SIGN	CRITE	ETA E	OR	AN	AI	ETI	ON	ολ	10	US	w	HE	EL.	СН	ATT	>					49
			ctives.																				49
			process																				50
	1.5		User C																				50
																							50
			Signal																				
			Closed																				51
			Autom																				51
			Path N																				52
			User In																				53
			Person																				54
		4.2.8	Miscell	aneou	ıs.																		55
	4.3	Hard:	ware																				55
		4.3.1	Basic I	esign	of	Cha	ir.																56
			Efficier																				58
	4.4		darity.																				58
			nsions.																				59
			lusions.																				60
						•	•	•	•	•	•	•	•	•	٠	•	٠	•	•	•	•	•	00
CHAPTER 5																							61
	5.1	Defic	iencies.																				61
		5.1.1	Aequisi	tion.																			61
			Human																				63
	5.2		nary																				65
									•		•	•		•	•	•	•		•	•	•		
APPENDIX	GL	OSSAI	RY OF '	PERN	1S																		73

ACKNOWLEDGEMENTS

I wish to thank my husband, Ricky, for all his assistance on this project. Without his love and support I would never have been able to complete it.

CHAPTER 1

INTRODUCTION

Twenty times a day an American, usually young and healthy, is involved in an accident which will leave him partially paralyzed for the rest of his life. Of the approximate seven thousand victims of spinal cord injuries per year, one thousand are rendered quadriplegic. Roughly twenty-five to thirty-five million people in the United States, ten to fifteen percent of the entire population, are considered permanently disabled. Of these, one third have impairments severe enough to curtail normal activity. The disabilities may be caused by congenital defects, accidents, war, degenerative nerve and muscle diseases and numerous other reasons. But, in spite of the cause, the need to restore independence, functionality and self-respect to these handicapped individuals is apparent. The advent of the microprocessor has the potential of solving many of the problems faced by the disabled community.

1.1 Evaluating The Need.

This athor believes that the importance of advanced technological aids for the handicapped could not he hetter expressed than by an individual in need of them. The following article was written by a man with amyotrophic lateral scherosis, hetter known as Lou Gebrit's disease.

Among the catalog of terrifying diseases one learns of at medical school, none frightened me more than amyotrophic lateral sclerosis. I have often asked myself why during the four years since I developed the illness

Was it the picture of that first patient etched indelihy in my memory? We had gathered in one of those awful, gloomy hasement lecture rooms that used to be an indispensable part of the architecture of teaching hospitals - rows of stark henches rising steeply from the "well", more appropriately the pit, which sported a blackhoard, a lectern and one medisval chair. That chair was occupied by a seemingly elderly man stripped to the waist. Perhaps it would he more accurate to say that the the chair occupied the man, so small, so cachectic, so vulnerable, so insignificant. He was reduced to flesh and bone, and what remained of his muscle was engaged in a macahre dance of death - the dreaded fasciculations of ALS.

His voice was all hut gone, and to this day, I wonder what private hell be silently endured during that hour as the neurologist detailed the myriad physical signs on his broken body. The hooming voice of the neurologist rendered the prognosis: "Hopeless! This is creeping paralysis! He will he demeaned, isolated, frustrated, unable to communicate, and probably will he dead in six months."

Today, I am the man in the wheelchair. However, I do not feel demeased. I have combated isolation by speaking out to those of my colleagues who still retain the obsolete views of my neurologist "reacher" and as a family we have tried to direct our frustrations into creative energy - all of which require the shilty to communicate. Host the use of my hands more than two years ago, and as I faced a track-colour, my wome nightmare was whether this would finally turn wow traffic.

Fortunately, miraculously, none of this would prove correct. On the day after my transhotomy, I received a letter from a fellow physician who also has ALS telling me what countless prior inquiries had failed to clicit, that a computer was available that could be operated with a single switch. Not fon the drawing board, not "just around the order of the country of

The significance of the single switch is that it can be operated by anyone, however physically handicapped, who retains the function of one muscle group. For me, my evehrow muscle is strong enough to depress a very light switch and thereby gives me access to the full power of the computer. Other patients have a functioning finger or the supplemental of the computer program had a switch fitted to her most hand bit down.

The screen presents the alphabet to me. A pointer ("cursor") moves at a speed that I select, and I press the switch when the pointer is next to the letter I want. This opens an electronic dictionary to a page of words heginain with the desired letter. The process is repeated, and with the aid of the pointer, a word is selected. It takes longer to explain than to do!

What has the computer, in coajunction with a printer and voices synthesizer, afforded me? I talk to my family -that is most wonderful. I can make conversation with friends - the jokes take a little longer, hut they don't seem to mind. I can work independently again - write papers, review manuscripts, cooperate by mail with other scientists and I am able to interact with the persons in my laboratory. I write out my ideas before we meet and sum up afterward. Because of the loyalty and devotion of this group - and for me the indispensable ability to communicate - our research continues to be original and productive.

After 27 years of marriage and total sharing, Pauline and I have discovered the joy and astalisation of writing together. Topies come from the news, from books we read, from our own experiences and from the buman comedy all around. Ideas seem to be generated simultaneously, one of us prepares a draft - I, of course, use the haste to use the rich textures of the Earlish hauetage.

A computer program such as mine, operating from a disketten ob higger than 45-rpm ercord, can change the lives of hundreds of patients who retain cognitive function but are analyte to communicate. While it was patients with arther and the removable of the patients with a repairing the analyte and in individual who have lest their vocal cords, should be equally helpful. The main problem appears to be a dearth of information among beathle-are problem appears to be a dearth of information among beathle-are and the second of the communication of the communication programs designed to meet the needs of the handicapped have appeared in the Summer 1983 issue of Communication Outlook; a publication of the International Society for Augustation and Alexandre Communication. However,

Coping with ALS is a grim ordeal - but it is not hopeless. Every patient needs an infrastructure of love and support. Now, for those of us who could no longer communicate, the computer provides ALS patients with an Alternative Life Style [MCWI84].

This article was written by Dr. David Rabin, a professor of obststrics and gracology at Vanderbilt University's School of Mediciae. Dr. Rabin was able to write this thoughtful essay through the use of computer technology and the movements of one cychrow. This is just an example of the bope offered to the disabled by the power, versatility and apeed of the computer.

Apart from the well-intentioned reasons, it makes good economical sense to restore independence to the disabled. In 1980, it was estimated that the average cost of spinal cord injuries was \$319,777 per year (forty-four percent direct medical costs, and fifty-six percent indirect income loss). Given that most spinal cord injuries occur when the victim is between eighteen and twenty-two years of age, the lifetime cost of the injury could accumulate to over fifteen million dollars if the individual should live to be sixty-five. Multiply this by the number of people with spinal cord lesions and the figures become astronomical [LEIF81]. The utilization of suitable technological aids could considerably reduce health care costs and even facilitate the handicapped individual's reinterduction into the ioh market.

1.2 Report Organization.

This document is concerned with computer aids for the handicapped. It is divided into five chapters. Chapter 2 is a review of the literature. It includes descriptions of hoth existing and potential aids, distinguishing hetween those which reduce the handicap and those which attempt to replace the impaired function. A review of wheelchairs, both manual and motorized, is presented in Chapter 3. This chapter also documents the history of the wheelchair and presents an explanation for the lack of innovation in its design. Chapter 4 proposes the specifications and design of an autonomous wheelchair. Chapter 5 concludes the report with a summary of the entire document. It also suggests problems with current methodologies and possibilities for future development. Finally, the appendix contains a glossary of terms.

CHAPTER 2

COMPUTER AIDS FOR THE HANDICAPPED

2.1 Introduction.

Terminology has been established by the World Health Organization to distinguish between three often missued expressions concerning the handicapped. An 'impairment' is a physiological deficiency, a 'disability' is the loss of function due to the impairment, and a 'handicap' is the social or economic cost of the disability [RAGGS3]. For example, damage to the spinal cord is an impairment, where paralysis resulting from the injury is the disability. Lack of mobility, reduction of employment potential, and loss of independence are a few of the consequent handicaps. This chapter presents an overview of technological axis for the handicapped, distinguishing between those which reduce the handicap or disability, and those which attempt to replace the lost function.

2.2 Reduction of Handicap.

The computer has long been considered a potential and for the handicapped, but until the advent of the microprocessor it was too large and expensive to be a viable alternative. When the microprocessor did become available to society, many bioengineers and rehabilitation experts recognized the vast opportunities afforded to disabled individuals. In the ensuing years, an extensive number of systems have been developed which serve to lessen the impediments of the handicapped person. An attempt to compile an intensive survey of all of these aids would produce an unvisidly document which could not do justice to the amount of time and effort spent developing the devices. This section, therefore, will present an overview of aids for the handicapped which should provide the reader with an overall view of the types of aids available and the handicaps they serve to overcome.

2.2.1 Vision Impairment.

Individuals suffering from vision impairments must face obstacles daily that sighted persons can not begin to appreciate. Deprivation of sight may well affect every superof a person's life. Simple things like taking a walk, reading a book, or attending a class can become difficult or even dangerous without benefit of vision. Rehabilitation experts have recognized that the microprocessor can alleviate many of the hardships faced by the blind. Aids have been developed which may assist the blind person in traversing a route, orienting himself with his surroundings, and even comprehending the printed word. The following sections highlight a few of these systems.

2.2.1.1 Mobility Aids.

One of the most difficult challenges for a blind person is simply getting from one place to another. Travelling unfamiliar routes can be a frightening and dangerous experience to the unsighted. Even frequently travelled routes, which may be anvigated from memory, can become hazardous due to transitory obstacles. The blind have found it necessary to rely on long canes, guide dogs and their other senses when travelling alone. These are all invaluable, but do little to warn a person of obstacles at or near head height not originating from the ground, such as overhanging branches. Studies have shown that a blind person travelling alone has a abover walking speed, must stop more often to orient himself with his surroundings, and has a higher heat rate than if he were travelling with a sighted companion [HEYESSA]. Scientists are currently developing instruments which should complement the established aids in supplying a blind person with enough information about his environment to allow him to procress with more confidence.

One such aid is the "Sonic Pathfinder," a mobility aid which utilizes ultrasound to warn the blind individual of obstacles in his path. The device evolved from a handheld system, which utilized ultrasound transducers to detect obstructions and played a series of musical notes to alert the user to them. Because ultrasound can not recognize drop-offs such as curbs or holes, the aid bad to be used in conjunction with a long cane or guide dog. In order to alleviate the individual from using both hands for navigation, the developers of the "sonic pathfinder" chose to mount the ultrasonic transducers on eyeglass frames. Ultrasonic receivers mounted on the right and left of the frames alert the individual to obstacles to his right or left, respectively, by signalling with a series of descending musical notes in the corresponding ear piece. A center mounted receiver warns of an obstruction directly in the user's path by signalling in both ear pieces simultaneously. Once the device has detected an obstruction in its direct path, it will focus only on that object until the user has navigated around it. This prevents the masking of a central obstruction by one to the right or left of the user. The design also allows the individual to utilize shore-line information so that he may maintain a constant distance from continuous barriers such as walls or hedges. A major concern during the design of this aid was that, because the user would need to his process additional information, his pace would actually be decreased. However, prototype tests have shown that the output is so easily understood that the user may maintain his normal pace while enjoying the increased confidence provided by the aid [HEYE83h].

Computer vision systems have been the hasis for other mobility aids for the blind. Representative aids typically process a picture of the environment and present it to the user through tactile stimulation. Several systems have been designed which essentially draw a two dimensional picture of the environment on the user's abdomen through the use of vibrators [COLL84], [KIRS85]. A microprocessor is utilized to examine the picture, extract key elements, and create appropriate signals to be delivered to the vibrators. It is left to the user to recognize objects through the stimulation and to take the proper action. Many experts helieve, however, that this type of device may actually hinder the user. They feel that the skin in the area of the abdomen is not sensitive enough to be able to handle the vast amount of information required. Another concern is that the user would need to concentrate so intently on the stimulation, that he would be impervious to other signals from the environment such as the sound of a car engine. To remedy these problems, another aid is being developed which places the hurden of object recognition on the computer. This device incorporates a pattern recognition system to isolate obstacles from the environment and alert the user to their presence through a speech synthesizer. The direction and proximity of each obstacle is presented to the user by tactile stimulation. In this case, stimulation is applied to the forehead through sixteen vihrators mounted on a headhand. When an obstruction is first sensed by the device, it is discretely reported to the user through earphones. The forehead stimulator most closely representing the alignment of the object is activated to convey direction, and its proximity is communicated by the frequency of the stimuli. Due to the nature of computer-vision, mobility aids, real time processing is essential. Prototype systems have been

developed which work well in limited test environments; however, state-of-the-art technology has not yet produced a device small and fast enough to be realistic in an everyday environment [COLL84].

Another type of mobility aid for the blind allows the user to locate or orient himself with respect to specific areas. The Sonic Orientation and Navigational Aid (SONA) consists of a transmitter which is carried by the user and receivers mounted at locations of inserts such as huilding entrances, exits, and restrooms. The transmitter features a keypad which allows the user to specify certain codes indicating the desired area. For instance, 911 might he the code for an emergency exit. If a signal is transmitted within seventy-five feet of a corresponding receiver, the receiver emits a musical tone, indicating the direction and proximity of the desired area. Popula acceptance of this device could provide a relatively simple and inexpensive method for hlind individuals to find their way even in unfamiliar surroundings (RKLL84).

The final mobility aid to be discussed here is a portable navigation device which hasically helps a person remember long or seldom travelled routes. The device allows the user to record the number of steps taken in various directions so that he will be forced to memorize such details. The aid operates in two modes. In the record mode, a pedometer counts the number of steps taken until the user indicates that he is changing direction by pressing the key associated with that change. At this point, the aid will record the turn, and the number of steps taken prior to that turn. When the user has reached his destination, he may store the entire route and associate it with a number for subsequent playback. Once a route is stored, it may be utilized in the playback mode to enable the user to once again travel that route. The device counts the aumher of strides taken by the user and emits specific tones to signify changes in direction. The playback mode may he used to either assist the user in travelling to a particular destination or to retrace his steps hack to the original starting position. In addition to turns, the user may record or be alerted to cross roads, pedestrian crossings and staircases [FREE84].

2.2.1.2 Reading and Writing Aids.

Perhaps the most difficult obstacle faced by the hind individual is isolation from the printed word. Many everyday rituals require the alhilty to read or write and this is especially true for vocational opportunities. It would he foolish to helieve that the hind could rely on braille for all their written communication needs. A 1976 survey revealed that only twenty-one percent of the visually impaired students enrolled in hoth schools for the hind and customary schools were proficient in braille [HOROS1]. Many hooks have never been translated to braille and those that have are more expensive and much more volunious than print. It has been estimated that the braille cell, together with its necessary encircling space, requires forty times the amount of space as ordinary print. This together with the thick paper needed for braille and the elevation of the dots themselves could make a dictionary large enough to fill an entire bookese (SULIASI).

Advances in technology have served to alleviate many of the written communication needs faced by the visually impaired. A variety of devices have heen developed which give the blind person access to computers, ordinary print, and braille stored on tape. Cost, finger sensitivity, and user motivation are the major factors in determining which device is hest suited for an individual.

2.2.1.2.1 Reading Machines.

The Kurzweil reading machine incorporates optical character recognition software and a speech synthesizer to give visually impaired or dyslexic persons access to ordinary print. The user simply places the printed page face down on the camera unit, and the derice automatically seams and vocalizes its content. The device is excellent for library use hecause the user requires very little training, but its relatively high cost makes it impractical for home use [KURZ81].

The Optacon, OPtical to TActile CONverter, is a device which discersa ordinary print and conveys it to the user through the sense of touch. The aid employs a hand held camera unit which the hind user must guide along the printed page. An array of vibrating pins under the index finger of the user's other hand mirrors the print image appearing before the camera lens. Utilization of the Optacon requires intensive training due to the natural difficulty a hind person would encounter attempting to scan something he can not see. It also demands an adequate degree of fingertip sensitivity so that the user may identify the images presented to him. Tests have heen developed which serve to ascertain an individual's shillty to use the device. Given the proper conditions and a substantial degree of user motivation, however, the Optacon gives the blind user access to a wide range of printed material. Separate lens attachments have heen developed which provide the user with access to compute terminal deplays. The currently planned additions of a speech synthesizer and automatic acanner will soon make the device substantially easier to use [SCHO81].

2.2.1.2.2 Transient Braille.

One of the major problems with hraille is its other volume. A significant improvement for the storage of hraille was the innovation of the transient braille display. This medium works on the same principle as the Optacon's display mechanism. Characters are represented through an array of pins which are elevated in the form of braille cells to convey meaning to the reader. The display is a mechanical device which inberently adds more expense and bigber risk of failure than electronic components. However, the usefulness of the mechanism may outweigh its disadvantages.

The advent of the transient braille display has offered new opportunities to the braille reader from which he was once excluded. Transient displays may be attached to computer terminals to give the blind individual full access to computer capabilities. Braille books may now be stored on cassette tape and played back through the use of special aids which employ the medium. Devices using this mechanism bave been developed which make it substantially easier for the blind student to take notes in class. The braille reader was once forced to enhose braille on tibick paper, a noisy and cumbersome procedure with no edit capabilities. Devices such as the Versabraille have been created which use the transient braille display in conjunction with a braille keyboard and tape recorder. The Versabraille allows the user to enter information onto tape which can later be edited and stored for future use. The instrument may also be utilized as an interface to the computer terminal, thereby allowing even greater flexibility DURRS4, [ROLLS1].

2.2.1.2.3 Speech Synthesis.

Many recently developed devices for the blind employ the technology of speech synthesis. The advantages offered the blind through this facility are obvious. Devices which incorporate speech synthesis are much easier and faster to use band devices relying on tactile stimulation. They also provide a reading facility to persons who for some reason are unable to use braille. Speech synthesizers employed by reading machines give the blind population access to printed material. Computer terminals utiliting the technology provide the blind individual with computer facilities. Exercising this technology for use by the visually impaired population does, however, have major disadvantages. Most aids employing synthesizers enable a person to read, but have no facility to allow him to write [MOYESI]. Also, few search or edit capabilities are available to individuals using speech synthesis as an interface to computers. Two significant problems are the limited vocabularies employed by many speech synthesizers and the fact that the untrained listener can very often simply not understand the output [ROSSSI]. [ZINGSI]. These problems have been overcome for more sophisticated synthesizers, but their expense is generally cost probibitive to the disabled population.

2.2.1.2.4 Braille Trainer.

Three of the five major causes of blindness effectuate a slow deterioration of the eyesight. The other two may either cause sudden blindness or may gradually render their victims sightless. Researchers have developed a device which makes use of this time to help an individual learn braille while he still retains some vision. The Computer Aided Braille Trainer, CABT, was designed to facilitate the learning of braille and to lessen training cost. Because many blind persons must attend special schools or bire private tutors in order to learn braille, the process can be both expensive and time consuming. Research has shown that being able to see a braille cell simplifies the learning process and assists an individual in memorizing its structure. Also, even a person with very limited vision is more likely to be able to recognize an object through sight than through an untrained sense of touch. The CABT comprises special bardware and software which interfaces with a COMMODORE 64 microcomputer to assist the user in learning braille. The keyboard is patterned after a Perkins Brailler, a braille embosser used by many blind individuals. When the user keys in a character, it is displayed in print on the top half of the screen and braille on the bottom half. The CABT is a simple, relatively

inexpensive device that could give an individual who is going blind invaluable practice in learning braille [SIBE84].

2.2.2 Hearing Impairment.

Computer technology has provided a variety of systems to aid the hearing impaired individual overcome his handicaps. Devices have been developed which allow deal people to communicate over telephone lines and to interpret sound through the sense of touch. Instructional systems have been implemented to provide the deaf with instant feedback when learning to speak or finger spell. This section provides an overview of typical aids for the hearing impaired.

2.2.2.1 Telecommunications.

Until 1984, the deaf had no access to a device which had become commosphere to most of society - the telephone. That year, a deaf physicist designed an account of society is the telephone that year, a deaf physicist designed an account of the telephone that the telephone times. This marked the beginning of the Telecommunications For The Deaf, TTY, network. The ensuing years saw a distinct standard being set in telecommunications for the deaf. The similarity of TTY components to those of computer terminals ensured a substantial price reduction when the latter was mass produced, enabling many disabled persons to purchase the device [WALS81]. Telecommunication devices utilizing current technology far surpass the limitations of the TTY; however, the established network is so firmly entreached in modern day society that there is little hope of changing the standard. Researchers, recognizing this problem, have sought to develop advanced aids which are either compatible with the TTY network or offer such significant improvements over the TTY that they may be considered visible alternatives.

Although the TTY is a very useful device, it has several deficiencies which further isolate deaf individuals from the hearing population. Its baud rate of 45.5 hits per second is very slow by today's standards. Its communications protocol is incompatible with the standard ASCII terminal. The TTY user is unable to store messages for subsequent transmission, and utilization of the device requires a similar aid at the other end of the communication link. Finally, it is a large and hully device, making it impractical for portable use. These factors combine to make the TTY useful for deaf individuals to communicate with other deaf persons or close friends and family, but otherwise unsuitable. Hearing persons without deaf friends or relatives are unlikely to ever own such a device. However, even given TTY access, a hearing individual, who is accustomed to instant communication via the telephone, is likely to find its utilization very slow and frustrating, leading to disuse. Compounding this, the slow band rate will necessitate higher charges for telephone line usage [LEVISt].

In an attempt to alleviate some of the problems associated with the TTY actwork, researchers have designed aids to both complement and replace it. Several devices have here developed which use DTMF (Touch-Touch signals as a communication medium. Utilization of this medium would require that special hardware he located only at the residence of the deaf individual. Anyone with access to a standard Touch-Tone telephone may communicate with deaf persons using the DTMF signals. Because there are three letters on each Touch-Tone key, various schemes have here developed to encode the signals to ensure against ambiguity. One such method involves the user depressing the key with the desired letter followed by entering either 1, 2, or 3, depending on the position of that letter on the key [JOHNS1]. Another scheme requires that the user depress the key with the desired letter one, two, or three times, again depending on the letter's position [SHEN94]. Characters from the

ASCII character set not appearing on the standard telephone keypad have been predefined to allow for utilization. A keypad overlay, available with most systems, allows for ease of use. Special equipment at the other end of the transmission link receives the DTMF signal and converts it to either the standard ASCII or TTY format. If the deaf individual is also mute, he would require special assistance in formulating a reply. This could be accomplished either through the use of Morse code, again using DTMF signals, prerecorded replys on tape, or the use of a speech synthesizer.

Another telecommunications aid takes advantage of the widespread popularity of the personal computer in today's society. The "Universal Translating Modern" was designed based on the premise that the personal computer is becoming a standard appliance in most bouseholds. The device allows the deaf to communicate with anyone owning either a TTY or a personal computer. Upon receipt of a transmission, the modern will convert between the ASCII and TTY character sets as necessary. This not only allows the desi individuals to communicate with a wider range of people, but also gives them access to a number of data bases currently available through telecommunications [BOZZ81].

2.2.2.2 Programmable Hearing Aids.

Microprocessor technology may well serve to reduce many of the problems currently associated with hearing aids. Hearing aids commonly employ a signal processing device to correct a distinct type of bearing disorder. However, there are many different types of audiological dyafunctions and each hearing impaired individual is likely to suffer from more than one of them. According to E. Villchur, "the effectiveness of signal processing in current hearing aids is probably greatly compromised by the lack of a systematic method of matching the processing, both

qualitatively and quantitatively, to the residual hearing of the person heing fitted [VILL81]." Unfortunately, hearing aid selection generally involves choosing from the available aids the one that most likely matches the hearing disorder. This crude method of selection could be eliminated by a programmable bearing aid. Ideally, the utilization of such an aid would allow for custom fitting for each individual's unique hearing impairment. Extensive testing must be performed, the proper signal processing pattern must be determined, and the hearing aid must be programmed to suit an individual's particular requirement. All of these functions would be greatly facilitated by the use of a computer. A programmable hearing aid should greatly reduce the need for special hardware requirements for hearing aids and alleviate many of the problems encountered by the hearing impaired individual VILL81.

2.2.2.3 Tactile Stimulation.

Persons who suffer from sensory deprivation must very often rely on their other sense for compensation. Researchers are attempting to optimize sensory substitution for the deaf through tactile stimulation. Several systems have already heen designed which discriminate critical components of speech and present them to the deaf user through withrators applied to the akin. Various systems utilize different signal processing techniques to analyze speech components; however, all systems work on the hasic premise that the deaf individual must learn to recognize meaning from withratory sensation. The systems generally consist of a microphone to capture sound, a microprocessor to analyze the acoustic signal and vibratory stimulators to propagate that signal. Prototype systems have been tested and have shown good results. Almost all persons tested were alle to understand complex sentences using a combination of vibratory sensation and lipreading. Further tests have shown this presenting all impressing skills improve considerably as a result of heing utilized in conjunction with

tactile stimulation. In fact, many persons tested were better able to lipread even when use of the stimulation was terminated [KANE81], [ROSO84].

2.2.2.4 Teaching Aids.

The computer's inherent tolerance for repetitive tasks deems it an excellent teaching facility for the disabled. A variety of instructional systems have been developed to aid the hearing impaired population in overcoming their handicaps. One such system aids the speech therapist in training a deaf individual to speak intelligibly. A child generally learns to speak through hearing his parents' voices and imitating their sounds. Parental response, along with the sound of the child's own voice, provides feedback necessary for the learning process. The deaf individual is denied this experience. The purpose of the automated speech therapy system is to offer visual feedback for deaf persons learning to monitor their own voices. Typical systems display graphical representations of critical speech components on a computer terminal. First, a graph representing the correct pronunciation of a word is created when either the therapist or a pre-recorded tape provides the correct enunciation. In a typical system, the correct graph is shown continuously on the ton half of the screen while graphs representing the student's pronunciation are displayed on the bottom half. The student may practice pronouncing the word and receive instant feedback on his progress [DUFF81]. Special speech therapy systems have also been developed for deaf children to help make learning fun. For instance, in "ball-game" the child must orally direct a pitch controlled basket ball along a predefined path in order to score a point. Another system for smaller children employs a "happy face" which teaches a child to monitor the volume of his voice by covering its ears when the child vells or speaks too loudly [HATO81]. Systems have also been developed to assist the deaf in learning to finger spell. These systems are typically comprised of visual displays representing the deaf alphabet, and they allow the user to practice recognition of words spelled out by an automated hand [KIHN81].

2.2.3 Motor Impairment.

The use of microprocessors has the potential of offering a great deal of independence to persons with motor disabilities. A vast number of systems have already been developed to provide communication and environmental control to those who are otherwise unable to manage these functions. Systems are also in development to provide entertainment and personal care facilities. This section presents an overview of these systems and possibilities for the future.

2 2 3 1 Communications

Man-machine interfaces have been developed to facilitate the use of personal computers by motor impaired individuals. This not only provides a means of written communication, but also if used in conjunction with a speech synthesizer, allows oral communication for the severely disabled.

The most opular interfacing method in use is a row/column scanning system. This method utilizes a two dimensional matrix of words and/or letters. The cursor scans down the rows until a switch is activated by the user, indicating that the current row contains the desired item. The cursor then scans through each item on the row until the user makes a selection. The selected item is placed on a line below the matrix and the scanning is resumed. Upon completion of the task, the user may select to store, print or vocalize the text.

There are several different theories concerning the optimal implementation of matrix scanning. Systems have been developed such that the items which are most frequently used are the most easily accessed. For example, the often used letters "e". "a", and "t" are positioned in the upper left hand corner of the matrix. This method significantly reduces selection time. Software has also heen developed to utilize dynamic matrices so that, hased upon the previous selection, items are positioned according to the probability that they will be selected next. For instance, the letter "u" has the highest probability of heing selected after the letter "q" and is positioned accordingly [JONESI]. Another school of thought holds that elements in the matrices should have a static position. Subscribers to this philosophy believe that selection times are increased with the utilization of dynamic matrices due to the time required for the user to orient himself to the new matrix. A system which uses static matrices allows the user to hecome familiar with character positions, thereby decreasing selection time WEISSS.

Other systems have been developed based on the inference that children, who have been unable to communicate for most of their liven, may be unable to form effective sentences. These systems allow frequently used phrases to be stored according to their grammatical use. For example, frequently used responses are stored in an "nawer" array, and typical questions are accessed from a "question" array. A special mode will allow the child to form unique sentences by prompting for information based on the rules of sentence systax. For instance, if the child has selected the phrase "will eat" from the verb array, the system will respond with the prompt "who" and display the noun matrix. A spell mode will allow the child to use words or phrases not found in the established arrays (TILLS!).

Rew/column scanning systems can be implemented using a variety of switching mechanisms depending upon the capabilities of the user. The minimum requirement is a single switch allowing a preson with only one functional muscle group to operate the system. Persons with greater muscular capability may utilize redundant switches in order to reduce fatigue, or may use enhanced scanning systems which take advantage of these canabilities.

Recently, row/column scanning software has been integrated with commercially available software packages. This integration precludes the need to develop special software packages for use by the disabled, thereby greatly enhancing the usefulness of the personal computer IKULI84!

2.2.3.2 Environmental Control Systems.

In order for the motor disabled person to achieve any independence, he must first be able to control his environment. Micro-computers can be used in conjunction with commercially available systems to allow the severely disabled a minimal but very innortant amount of environmental control.

Typical environmental control systems allow the user to manipulate a number of electrical appliances by means of a microprocessor. A list of controlled devices may be accessed by a disabled individual through the use of a voice recognition unit or scanning software such as that described above. The user selects a particular mechanism and indicates that it be turned on or off. The system may be enhanced to operate drapes, doors, and similar objects if connected to a motor driven device which can be regulated by the controller. Some systems allow the user access to the outside world by means of a standard telephone equipped with an external speaker. The user may dial by selecting the proper sequence of numbers and issuing the dial command.

Many quadriplegies are confined to their beds for the majority of each day. An environmental control system using a voice recognition unit was recently developed to enable a patient to utilize the system from either his bed or his wheelchair. The Voice-Operated AC Controller (VOAC) was built using a X-8 Zijioc CPU and a voice recognition integrated circuit. The system utilizes an LED display on which appliance and command codes are scrolled. There is no display until the user activates the scrolling process by speaking the word "SEARCH." The commands "FASTER" or "SLOWER" may be used to alter scrolling speed. When the proper appliance and code are displayed, the user terminates scrolling with the "STOP" command. Upon receiving the "GO AHEAD" command, the system generates the proper signal and transmits it along the house wiring. The signal is decoded by a commercially available carrièr-current-operated switch at each appliance outlet to determine if an action is required. Upon receipt of the proper signal, power is either supplied or terminated to the anoliance HOGASSI.

2.2.3.3 Robotic Aids.

Robotic aids have the potential of providing many manipulative functions required to restore the independence of the quadriplegic. Robotics have been used in industry for some time to perform repetitive functions in a controlled earironment. Recently, rehabilitative engineers have applied this technology to aid the handicapped. Typical aids are comprised of telemanipulative functions, which are completely controlled by the user, and robotic functions which are controlled by a computer. Telemanipulators offer full flexibility of function but are alow and eumbersome to see because each step of each function must be commanded. Computer control allows repetitive functions such as feeding to be executed quickly and accurately.

Researchers at the Johns Hopkins University Applied Physics Laboratory have developed a robotic arm which when used in conjunction with a structured worktable provides some manipulative capabilities for the quadriplegic. The system provides the user with the facilities necessary for reading, eating, and using a telephone, environmental controller, or personal computer. The robotic arm/worktable and the user's wheelchair can be operated by the same low profile cbin controller. An infrared link provides the control path between the wheelchair and the robotic arm, eliminating the need for special assistance in using the system.

The robotic arm is preprogrammed to perform specific functions. The user selects the desired program from the LED display which scans the available preprogrammed tasks. These programs require that the work table be structured in a specific manner, for example, the program "PICK UP BOOK" demands that the book rack and reading stand be in specific locations. Similarly, the patient may operate the telephone, environmental controller, or self-feeding facilities. User specific programs may also be written using a specially adapted forty key calculator keypad which can be connected directly to the micro-controller of the robot arm.

The user may also select to operate the system in manual mode using the arm as a telemanipulator. In this mode, the patient may select one of the six available degrees of freedom from a scanning display by issuing a command through the use of a discrete switch such as a puff sensor. He may then control the selected joint by manipulating the chie controller. Movement of the joint is proportional to displacement of the controller SCHNSII.

Researchers at Stanford University are developing a mobile robotic aid for the handicapped. The aid consists of a commercially available industrial robot arm mounted on a three wheel omni-directional base. The user may elect to employ preprogrammed manipulative tasks or may assume direct control through a voice recognition system. A small camera mounted on the robot's arm provides a video signal which is transmitted to the console, allowing the patient to monitor and control its movements even from another room. The robot is equipped with a bumper seasor

to protect it from unexpected objects in its path. If the robot does encounter such an object, it will back up, move to the side and attempt to proceed, in effect, moving around the object.

Voice control of the robotic aid proves to be difficult to implement due to the fact that the English haguage is task rather than movement oriented. Currently, in the prototype system, researchers use a very small portion of the Kurzwell Voicesystem's 1000-word recognition capability. The present command language is highly technical and would be cumbersome for the average user. A natural command language is planned for the final system, and will require at least two additional years to implement SLAT986. [LEUS1].

Currently, motors for robotic arms are large, heavy, and subject to mechanical failure.

Research is presently underway, however, to replace these motors with lightweight, more compact, and more reliable drive systems. Researchers at Duke University are developing a robotic arm which, because of its few moving parts, is less prone to failure. The arm is comprised of flexible tubing which is corrugated on one side. Computer controlled air blasts cause the corrugated side to inflate, thereby bending the arm in the opposite direction. This design allows the arm to bend at any point.

Presently, it only bends in one direction; however, researchers believe that the use of spirally corrugated tubing will increase its flexibility [TYND87]. Another innovation currently being studied by researchers in the shape memory alloy (StA)s effect as an actuator for robotic arms. According to Dr. L. Schetty, SMA is based on the principle that "a part deformed or stretched at one temperature will, upon being heated to a second temperature, completely recover its original shape [ScilEss]." Hitschi has applied this technology to create a compact, lightweight robotic hand. Two spring louded fingers and a thumb are regulated by pulling their associated drive wires. SMA

strips contained in the forearm control the tension on these wires by contracting when electrically heated. The size and weight of the entire robotic arm is reduced by eliminating the need for electric motors [NAKA84].

2.2.3.4 Entertainment.

The need for a severely disabled person to communicate or to control his environment is certainly very great, but probably more important for the disabled child is the need for entertainment. Generally, by the time a child is given the power to communicate through a computer, be has already adapted to his lack of verbal exchange. Also, regardless of his intellectual ability, the non-oral child is commonly spoken to less often and less intelligently than if he were normal. Very often, by the time a child is exposed to the computer, he has lost any incentive to attain the slow communicative abilities it can provide. The need for pleasurable activities that take advantage of a child's remaining motor abilities is apparent. The child may not only express distributions that the computer is a potential and [COLDS1].

E.P. Goldenberg of Tufu University School of Medicine has recognized the need for computer systems which entertain as well as teach a child. Utilizing the knowledge that many motor disabilities are characterized by lack of control rather than paralysis, Dr. Goldenberg has developed an algorithm which filters the "nouse" from an input signal generated by victims of certain diseases. He has implemented this process to enable athetoid children to use a joystick or graphics tablet to play computer games. The graphics tablet also lends itself to freehand drawing or writing which gives the child a much welcomed opportunity to produce recognizable figures. Computer interaction has served to replace some of the everyday learning experiences denied to the handicapped child. For example, one bright fourteen year old demonstrated an unawareness of the behavior of a bouncing ball when playing a computerized version of ping pong for the first time. After a period of time, the child apparently began to understand, when he was able to anticipate the action of the ball [COLDS1].

It is also very important for the severely dishibled adult to have recreational outlets.

Much research and development has been done on communication and environmental control systems, however, entertainment facilities have been somewhat neglected.

Some systems have been developed to allow a user to express himself through musical composition and art; however, much more work needs to be done.

2.2.3.5 Summary.

It is important that many of the facilities available to the motor disabled be incorporated into an integrated system. One such system was designed specifically for an eighteen year old man who was left quadriplegic, non-vocal and nearly blind as the result of an automobile accident. The system provides a combination of communication, eavironmental control, and entertainment facilities. The user may access any of the features through a series of top-down menus which are displayed on the acreen as well as being vocalized through a speech synthesizer. The user may control mean selections either by noise activation, a sip and puff switch, or a touch switch that can be controlled by any functional part of the body. The system offers communication by allowing the user to build sentences from stored phrases, words or letters, which he may then store, print or vocalize using a speech synthesizer. Appliances can be managed through the use of an environmental control mode. The system also allows a person to express his individuality by creating music and color graphics. The music is created according to user response to special means, and may be stored and played back on a music synthesizer. Color graphics are created in

conjunction with the music. The time and effort devoted to creating this system had to be considered worthwhile when, after a short training period, the patient was able to form the word "APPRECIATE" [HOLMS1].

2.3 Replacement of Function.

The eyes, cars, and other sensory organs transform external stimuli into an electrical signal which can be interpreted by the brain. The brain reacts by transmitting signal to muscles which respond with the appropriate movement. Any interruption of the communication will result in the loss of sensory or motor function. The technology of applied neural control offers the possibility of replacing those functions lost due to physiological impairments [MORTSS].

Applied seural control utilizes the fact that nerve fibers and muscles can be electrically stimulated to produce a desired action. Electrical stimulation has enabled functionally blind persons to see light, deaf persons to hear sound, and paraplegies to walk. Practical utilization of the technology is years away; however, continued research offers hope.

Electrical stimulation has been practiced in the medical field for some time. The Roman physician, Scribonius Largus first observed the healing powers of electricity in 40 A. D. when a patient was curred of gout after stepping on a torpedo ray, a fish which produces an electric shock. Later, when scientists first learned to guarate and store electricity in 1740, curse were reported for paralysis, kidney stones, epilepsy, and other disorders. Benjamin Frahlin is reported to have had good results relieving convakious with electrical shock (CHIZSSI). As important milestone for the technology was reached in 1927, when it was discovered that electrical stimulation could control the heart rate, leading to the development of the cardiac pacemaker [HFTCSS].

In recent years, more emphasis has been placed on the use of electrical stimulation in medicine. Methods are being developed to treat scoliosis (curvature of the spine), incontinent bladders, respiratory dysfunctions and spasticity MORTS3]. Computer chips are being designed to implant in the body to not only reconnect damaged nerves but also to restore their functionality [EDWAS7]. Research is also underway to restore sight for the blind, bearing for the deaf, and mobility for the paralyzed.

2.3.1 Visual Prosthesis.

Scientists are attempting to restore eight through visual providesis. It has been known for some time that electrical stimulation of the visual cortex of the brain will tasses a person to see a spot of light, a "phosphese". Furthermore, it has been shown that a direct mapping exists between the visual cortex and the retina of the eye. This knowledge has been the basis for several experiments in which electrodes are implanted or inserted percutaneously onto the visual cortex. The tests have been largely unsuccessful from the patients' point of view, but have produced valuable information for researchers.

The first visual prosthesis was developed and implanted by GS. Brindley in 1967.

The implant consisted of two silicone rubber caps, an extracranial cap enclosed between the scalp and the skull and an intracranial cap placed beneath the skull at the visual cortex. Eighty radio receivers attached to the extracranial cap were

connected to eighty electrodes attached to the intracranial cap via wires extending through a hole in the skull. Control of the electrodes was established by radio signals generated from outside the hody.

There were two major problems with this experiment. The design of the prosthesis required that the patient wear a belmet containing radio transmitters monated directly opposite their respective receivers on the implant. It was very difficult at that time, however, to make radio transmitters both small and powerful enough to fit the required number onto the helmet. Another problem encountered was that the number of radio channels suffered a high failure rate over time. Although the proathesis was made of the hest materials available at that time, the high voltage and compact size required in such a device made it susceptible to failure. The proathesis was removed in 1882, fifteen years after it was implanted. Of six receivers cut out of the extracranial cap, only one remained functional DONASSI.

The second visual prosthesis was created by P.E.K. Deaaldson and E. Sayer and implanted in 1972. The havic design was similar to the first prosthesis, but technology had yielded smaller and more adequate materials for implantation by that time. One major difference was the use of "AND" logic in the transmitting and receiving of signals. This time five "row receivers" and fifteen "column receivers" controlled seventy-five electrodes. In this method, two receivers must be activated simultaneously in order to apply voltage through an electrode. The use of "AND" logic, therefore, significantly reduced the number of radio transmitters and receivers required in the proothesis.

Once again, the prosthesis proved more valuable to researchers than to the patient. Induced phosphenes appeared large and indistinct to this patient, where the first patient had discerned small, bright shapes. Scientists believe that this was due to the extensive amount of time the patient had here hlind. In spite of this obstacle, researchers were able to elicit phosphenes in the form of hraille symbols which were recognizable to the patient. He was able to distinguish ninety percent of the symbols when shown them for only a second. Given an unlimited time period, the patient could discern all of the symbols.

This second attempt at visual prosthesis fell to the same fate as the first. The number of induced phosphenes decreased with time, presumably due to corrosion or failure of implant components [DONASS].

Other attempts at visual prosthesis have subsequently hera made. Each attempt has taken full advantage of technological advances, but much progress is needed befor the process can he of any practical use. Components must be small enough to fit in the available space in the head, yet able to operate at the relatively high voltages required in a visual prosthesis. Logic must be devised to produce the hest possible use of these components, and each element of the prosthesis must be encapsulated to protect it from the hard environment of the hody DONASSI.

Although attempts at visual prosthesis have heen largely impractical so far, continued research could be very valuable. Among the possibilities for this exciting field, is the thoory that it might someday he utilized to above a hind person scenes from his own environment. If the problems with a visual prostlessis are overcome, and advances are made in computer vision technology, it is reasonable to believe that a camera could transmit a picture to a computer which would then filter "noise" from the scene and generate signals to be transmitted to the prothesis. The scene could then he recreated in the form of phospheses to be viewed by the patient. As research

continues, scientists will increase their knowledge of the biomechanics of vision with an ultimate roal of the restoration of sight.

2.3.2 Cochlear Prosthesis.

Scientists are developing a method to restore some auditory sensation for the deaf through a cocalient proutbesis. In order to comprehend the function of such a device, one must understand how the ear responds to sound waves. Changes in the atmospheric pressure between 20 and 20,000 Hz. enter the external ear causing a corresponding movement of the ear drum. The ear drum along with the three small bones in the middle ear transform sound into mechanical energy. The vibration produces fluid movement in the inner ear, the cochlea, which causes hair cells on the basilar membrane to vibrate, exciting the neurons at the base of each cell. The neurons produce an electrical signal which is passed along the 30,000 fiber auditory nerve to the brain. It is damage to the hair cells of the cochlea which is the major cause of declaress PARKSSI.

State of the art medical techniques have made it possible to repair almost any disorder of the external or middle car; however, this is not true for the inner car [WHITS1]. A cochlear prosthesis is an artificial device designed to replace the lost function of the inner ear. This device must transform sound into electrical stimulation which can be interpreted by the brain. To accomplish this, the preschesis must consist of three basic subsystems: 1) an external transmitter/signal processor unit, 2) internal receiver electronics, and 3) an electrode array. The external device is responsible for collecting sound, transforming it into the proper stimulation pattern, and transmitting it to the internal receiver. The receiver, in turn, transfers the signals to the proper electrode(s), which delivers the electrical impulse to the auditory nerve.

Electrode arrays can be configured in several different manners. The first generation of prostheese was implemented using a single channel configuration, one active electrode and one return electrode. This provided the user with sufficient auditory sensation to distinguish environmental sounds such as the telephone ringing, a siren, or a knock on the door. The technique also enhanced a pasient's lip-reading shifty and allowed him to monitor the volume of his own voice [PARKS3]. Subsequent research has proven that multichannel stimulation has increased the patient's speech discrimination capabilities. According to C.W. Parkins, "there is now good evidence that a multichannel information system can produce significantly better speech discrimination than a single-channel information systems" [PARKS3]. Therefore, in the future, it is more likely that a multichannel system will provide a patient with not only speech discrimination but also speech intelligibility.

There is some dehate as to the optimal placement of electrodes in the inner ear. Experiments have been conducted such that electrodes are positioned directly on the auditory nerve. This has the advantage of requiring lower stimuloss levels than are needed by electrodes placed in the cochlea. However, the potential damage to the auditory nerve greatly outweighs the benefit of low power [PARKS3]. Another school of thought holds that there should be no invasion of the cochlea whatsover. Subscribers of this theory believe that insertion of electrodes into the cochlea risks further damage to neural structures. Their approach is to place the electrode on the promontory, the round window at the entrance to the cochlea, and optimize its use through single channel signal processing. The minor surgery required of this method allows the patient to obtain a temporary implant, so that he may assess the usefulness of the device without making a long term commitment. This method has only been implemented using a single pair of electrodes; therefore, the major disadvantage like with the problems inherent with a single channel device [ROSES1]. The final and most widely used approach is the insertion of the electrode(s) directly into the ecochies. New methods of insertion have minimized the threat of neural damage. Selective placement of multiple electrodes along the basilar membrane enable the stimulation of targeted neurons, taking full advantage of multichannel signal processing [PARKS1].

The external processor is probably the most complex of the three components of a prosthesis. It must function to collect sound waves, convert them into the proper electrical signal and transmit that signal to the internal receiver. Some researchers believe that the entire acoustic signal should be transmitted to the brain; however, most believe that the signal should undergo a feature extraction process. State of the art prostheses consists of approximately eight to ten data channels, but the ear has close to thirty thousand. According to R.L. White, "In this restricted-bandwidth system it is imperative that the few channels available be used effectively. The limited bandwidth must be utilized to transmit only the significant features of speech" [WHITS1]. In this case, the external processor must employ a feature extraction algorithm. In some multichannel systems, the external unit is also utilized to separate the signal into frequency bands so that each distinct band can be delivered to the appropriate electrode.

Deaf people are very susceptible in life threatening situations because they can not bear warning signals that the hearing take for granted. They must endure the everyday inconveniences of being unable to hear the door bell or the sound of a baby crying. But worst of all, they must suffer the social isolation of silence, of being unable to communicate at will. As technology advances, cochlear prontheses should alleviate many of these problems. However, before the prootheses can supply practical treatments for deafness, several obstacles must be overcome. Coding schemes for effectively processing an acoustic signal must be perfected if speech intelligibility is to be attained. The electrodes, conductors and internal receiver will require further miniaturization and specialization if the number of data changes currently used in found to be inndequate. If these tasks can be accomplished, the cocklear prosthesis should alleviate many problems for a large portion of the deaf population.

2.3.3 Functional Neuromuscular Stimulation.

made.

Approximately twenty people are partially paralyzed every day due to an accident involving damage to the spinal cord. Depending on the location of the injury, the patient may be paralyzed from the waist down (paraplegic) or from the neck or shoulders down (quadriplegic). In either case, be loses not only the use of his limbs, but also a great deal of independence, job potential, and in some cases, tragically, even his self-respect. Researchers are developing methods of belping these victims by restoring some of their host functions through the technology of applied nearal control. Functional neuromuscular stimulation (FNS) is the practice of electrically exciting muscles to restore some use to limbs paralyzed due to nerve damage. Paralysis is generally caused by an injury to the spinal cord, the transmission link from the brain to other parts of the body. In most cases, paralyzed limbs are isolated from any controlling signals from the brain, but are otherwise undamaged. It is the aim of scientists working with FNS, to bypass the spinal cord lesion and deliver appropriate signals to muscles in order to restore their functionality. The task is very difficult due to the complexities of the central nervous system, but significant progress has been

Researchers have been studying the possibilities of how electrical stimulation of muscles might aid paraplegies and quadriplegies for over twenty-five years. In 1960, a paraplegic was able to stand for several minutes when his leg muscles were electrically stimulated. Electrodes attached to the skin caused muscle contractions in his legs, which, in turn, straightened his knees. Since then, paraplegics have been able to stand, walk short distances, climb stairs, and even ride bicycles. Quadriplegics have been able to attain enough hand and arm control to he able to feed themselves and brush their teeth. The future of FNS holds a great deal of promise, but many problems must be overcome before the system can be made commercially available CRISESS.

One of the most hasie difficulties faced by researchers is determining the optimal method of muscle stimulation. For example, surface electrodes attached to the skin are clearly the simplest and most convenient means of stimulation, however, because the electrical impulse must travel through the skin and fatty tissue of the body, little selectivity is achieved and a number of different muscles and nerves are stimulated. Surface electrodes also suffer from slippage during use and many must be replaced daily. Intramuscular electrodes inserted through the skin offer much more selectivity, but their installation, while not requiring surgery, is still very difficult to perform. Another disadvantage is the fact that anything inserted through the skin eventually leads to infection and is rejected by the body. The hest alternative for long term use appears to be the implantation of electrodes directly on specific nerves. This, however, does require surgery and is therefore impractical to use during research [CHIZSS].

Another factor constraining FNS research is the extreme complexity of the human motor control system. Each muscle unit consists of thousands of muscle fibers and one motor nerve. In a healthy body, the brain generates a signal which is transmitted through the spinal cord to a motor nerve. The muscle fibers, stimulated by the motor nerve, convert chemical energy in the hody into a contractive force causing movement. The brain must coordinate and transmit an enormous number of signals simultaneously in order to create natural movement. FNS researchers have been able to establish redimentary control of paralyzed limbs, but they are nowhere near simulating the natural movement produced by a healthy hody. According to H.J. Chireck, "We simply don't yet know how to coordinate the transmission of signals to and from a large number of electrodes simultaneously. No more than 50 electrodes are now heing used even in the most sophisticated experimental systems. The hody's natural system of motor control, in contrast, employs thousands of simultaneous signals to regulate movement: (CHIZSS)"

To date, most FNS systems are controlled by one of two methods. The first technique, most frequently used for paraplegics, is controlled by a hinary switch. When the switch is turned oa, a computer generates the signals which activate the proper electrodes to came movement. An individual might walk, climb stairs or ride a hicycle. In any case, the computer must be preprogrammed to generate the proper signals for the desired movement. The method is satisfactory for many of the repetitive motions needed by paraplegies for activities like walking, but offer very little help for quadriplegies. The second controlling technique utilized by FNS systems is much more suitable for quadriplegies. This method involves an individual using a muscle or muscles sunficted by the injury to convey the desired movement. For instance, a quadriplegic might more his right shoulder to indicate that be wishes to extend his right arm. Electrical impulses causing the shoulder movement can be detected and subsequently analyzed by a microprocessor which would then determine the correct stimulus patters to extend the right arm. This technique is obviously much better suited for the non-repetitive movements needed for arm and hand control movements are for the non-repetitive movements needed for arm and hand control

required to aid the quadriplegic, but it also offers distinct advantages to the paraplegic. Researchers are currently developing an FNS system for paraplegics which allows the upper body (above the spinal cord lesion) posture to control the patients' movement. Their theory is that the system is much safer than the binary switching method because posture-mapping can determine if the patient's balance is adequate for the execution of a step and thereby prevent potentially hazardous movements (GRAUSS). Another system in development allows a patient to control his gait by simulating the ambulatory motion with two fingers [FHOMSS].

In spite of the progress in recent years, all PNS systems are experimental and will not be commercially available until many problems are solved. At present, all paraplegies using FNS in laboratories require auxiliary support to maintain balance. Many researchers believe that the development of a practical system, even for use with crutches, or walkers, will not be completed for at least a decade. A system not requiring the use of a support device is at least twenty to thirty years further in the future. According to M.J. Rosen, "few experts have been bold enough to predict that a system with the basic capabilities of moving a leg to catch one's balance can be developed for practical use within their lifetime. [ROSESS]"

When FNS systems are commercially available, quadriplegies are much more likely to benefit from them than paraplegies. Paraplegies are already afforded a great deal of mobility through the use of wheelshairs. FNS systems would need to increase mobility significantly before they would become a legitimate financial alternative. The system would also need to be easy to use and inconspicuous before many paraplegies would use them. For instance, braces are available for use by many paraplegies, however, the added capabilities they provide often do not warrant the physical and meatal effort they require. Quadriplegies, however, are much more likely to find the systems invaluable. Adequate control of the arms and hands will afford them some independence and eliminate the need for full time nursing care. This alone could save millions of dellars in health care costs; therefore, insurance companies are likely to provide financial assistance.

2.3.4 Computer Utilization.

The practice of applied neural control would be unrealistic without the use of computers. Computers have played a major role in the investigation of biomechanics to determine the optimal method of employing the technology. They are used during testing of prototype systems to record and analyre results, thereby allowing further improvement of the system. In the case of visual and cochlear prostheses, computers must perform signal processing on data gathered from the environment. Another important function is the determination and generation of the proper stimulus pattern to elicit the desired response. In the future, computers will also be used to analyze feedback data in order to insure proper operation of the system and to protect the body from any injury caused by the system.

2.3.5 Conclusion.

Significant progress must be made before the application of neural control can be perfected. Internal components must be ministarized and readered biologically inert so that long term subcutaneous devices are practical. External components must also be ministarized and packaged inconspicuously to avoid drawing unwanted attention. As knowledge of the biomechanics of the body increases, the effectiveness of these systems will also increase.

CHAPTER 3

WHEELCHAIRS

3.1 Introduction.

This chapter presents a study of wheelchairs, both manual and motorised. A description of the evolution of wheelchairs is included along with a discussion of the problems resulting in the lack of innovation in wheelchair design. This chapter is intended as a background study for the design of an autonomous wheelchair.

3.2 History of The Wheelchair.

The mobile indoor vehicle has a long, poorly documented history. Researchers have been forced to glean information about the carliest such vehicles from the drawings and engravings of relies from the past. Such evidence has included illustrations on an louic wase dated approximately 530 B.C., engravings on two Chinese stone coffins from the 6th century A.D., and woodcuts from the Middle Ages. This information indicates that mobile indoor vehicles were in use long before there was any written documentation attesting to that fact.

It was not smill the 15th century, that detailed descriptions of the predecessor to the modern day wheelebair were recorded. The most notable vehicle of that time was a mobile chair designed and built for Philip II, the King of Spain. This chair bore an annaing resemblance to wheelchairs built three centuries later. The chair was equipped with four small wheels and featured a redlining back and elevated footrests. The major difference between this chair and the subsequent wheelchair was the absence of any means of self-propolation. The first self-propelling wheelchairs were built in the 17th century. The preliminary chairs of this type were driven by two cranks which turned a single front wheel. The first chairs to be propelled by rotating the two large wheels were built almost a century later. It was not until the Givil War that the wheelchair came into use in the United States. At that time the wheels were made of wood, and it was not until the bicycle crans of the late 1800's that wire-spoked wheels were adopted for wheelchair use. As with all wars, World War I stimulated the demand for assistive medical devices, and the need for wheelchairs was intensified. The increasing popularity of the automobile prompted the demand for easy handling chairs which were light and foldable. From here, the wheelchair evolved into the standard wheelchair of the present iKAMESS.

3.3 Present-Day Wheelchairs.

The wheelchair is not just a means of transportation. Through the use of a wheelchair, persons who would otherwise be confined to bed are provided the freedom to move about the home, office, or even the outdoors. It is a practical aid which may preserve the heart, sustain the body and motivate the will. For many sick or disabled persons it can mean the difference between dependence and independence.

There are several different types of wheelebairs currently available to accommodate the various needs of the disabled community. This section presents a description of each type.

3.3.1 Nonfolding Chairs.

Still in extensive use today is the traditional, bigb-backed, institutional chair. Its inexpensive, sturdy construction makes it an ideal choice for use in hospitals and nursing homes. The chair is generally made of wood with caned backs and seats. Some models feature reclining hacks and elevated footrests. The chair is generally quite large, but pillows may he used to compensate for the had fit. Modern day, nonfolding chairs are built from tithular steel and are much smaller than the wooden chairs. This type of chair is generally used for transporting patients in hospitals and nursing homes or to provide a patient with a change of scenery during long periods of butters.

3.3.2 Rearwheel-Drive Chairs.

The rearwheel-drive chair is the most frequently used wheelchair and is the most appropriate for the majority of wheelchair users. Commonly known as the standard wheelchair, it is well-suited for hoth indoor and outdoor use. The chair is customarily made of steel tuking with the hack and seat composed of fabric or plastic. A user with adequate strength in his arms and hands can propel himself by terming the two large wheels positioned at the rear of the chair. Skilled users can tilt the chair hackwards, raising the front wheels enough to climb curbs. Push handles are located at the upper back of the chair for use by an attendant. A tipping lever positioned at the lower hack of the chair provides easy tilting by the attendant for navigating curbs. The chair usually weighs at least forty pounds, however, there are more expensive models which are lightweight but less stordy.

3.3.3 Sports Chairs.

The advent of a lightweight, sturdy wheelchair used in sports could be one of the most significant advances in wheelchair technology in recent years. Sports chairs have only been available on the commercial market for a short time, but wheelchair athletes have already tested their strength and versatility [BRUB86]. The sports chair features a solid frame constructed with a lightweight, corrosion-free alloy which makes it much more robust than the standard chair. Wheels on the sports chair may be attached at different levels to provide optimal performance by the athlete [KAMES6].

It is currently the policy of many physicians to prescribe a standard wheekhair unless justification is given to acquire a specially-equipped chair such as the sports chair. However, the performance and durability of the sports wheekhair is so commendable that it has been suggested that this policy be revened. Let the physician prescribe a sports wheekhair unless justification is given to prescribe a standard chair. Unfortunately, the policies of many third-party reimbursers have prohibited the general acceptance of the sports chair. One can only hope that the technology stillined in the creation of the sports chair be adapted in the design of other wheekhairs IRRITENSI.

3.3.4 Frontwheel-Drive Chairs.

The frontwheel drive chair is mainly for indoor use. With the two large wheels located at the front of this chair, its major advantage is that it requires less space than the standard chair, thereby making it much easier to maneuver in tight spaces. The location of the wheels doos, however, make lateral transfer in and out of the chair impossible. This chair is less likely to tip over backwards than the standard chair making it more suitable for users without legs or legs which have atrophird. Because the user must lean forward slightly to propel himself in this chair, he must have good treak stability to avoid falling forward.

Though primarily for indoor use, the chair may be used outdoors if curbs can be avoided. The location of the wheels makes it extremely difficult to negotiate curbs even with the aid of an attendant. The front wheel drive does, however, smoothen the ride over uneven terrain.

3.3.5 Amputee Chairs.

The ampute chair is a wheelchair designed specifically for the user without legs or with atrophic legs. Wheelchair users with either of these conditions may have difficulty with a standard chair due to the instability created by the lack of weight at the front of the chair. For instance, when travelling up a steep incline, the user's weight is positioned predominately at the back of the chair. The absence of weight at the front could easily tip the wheelchair backwards. The wheels of the amputee chair are repositioned to compensate for this unusual weight distribution.

3.3.6 One-Arm-Drive Chairs.

The one-arm-drive chair is a wheelchair designed specifically for users with one arm missing or paralyzed. The chair features two handrims on one side of the chair. The first is the typical handrim attached directly to the adjacent wheel. The second is positioned outside the first, and is connected to the wheel on the other side by an axle. The outer handrim is smaller than the inner one to simplify the use of one bandrim at a time. The double handrim may be fitted to various types and sizes of wheelchairs. The major disadvantage to this chair is that self-propulsion can be difficult to learn for those who are slightly disoriented.

The one-arm-drive chair was specifically intended for wheelchair users with the use of only one arm; however, others may benefit from the design. Users with a weak upper arm may tend to avoid using that arm when propelling a standard wheelchair. The use of a one-arm-drive chair, even for a few weeks, would force him to exercise that arm. The chair may also be well-suited for persons who need one arm free for carrying tbings. For instance, a mother might prefer this type of chair so that she can hold a baby in her free arm.

3.3.7 Motorized Chairs.

Motorized wheelchairs are available for use by the severely physically disabled or those who must conserve energy due to an illness. Motorized chairs are usually operated by displacing a control knob or joystick in the desired direction of travel. The chair is recommended for use by individuals who cannot otherwise operate a wheelchair, or for persons with cardine problems or poor museular endurance [KAME88]. Until recently, the use of a motorized wheelchair was not recommended in many cases because it was felt that the user would benefit from the exercise expended in self-propulsion. However, there is now a growing conviction among health care professionals that the wheelchair user should conserve his physical energy for useful tasks other than self-propolement [SEAMS1].

Special control mechanisms are now available for persons with very limited musels function. Among them are voice control through the use of a voice recognition system (MLLS8), and chin control indicated by the directional displacement of the chin (SEAM79). Also in use is a sip-and-pulf mechanism which requires the user to suck or blow through a mouthpiece(s) to indicate direction and velocity [KAME86]. A wheelebair is in design which allows the user to direct his chair simply by turning bits head in the desired direction. The position of the head is determined through the use of ultrasonic seasons mounted on the back of the wheelebair [JAFF81]. Still in the experimental stage is a myoelectric control mechanism. This device is able to determine a muscle's activity by measuring the electrical impulses atimulating that muscle. Through the use of this mechanism, the user could control bis chair by flexing different nuscles to indicate the velocity and the various directions. This type flexing different nuscles to indicate the velocity and the various directions.

of device could also be utilized in conjunction with a scanning display so that the user could select the desired command simply by flexing a single muscle [ROSS78], [DUR178].

Some of the disadvantages associated with the use of a motorized wheelchair are that spastic movements or movements made during periods of inattention are amplified [KAME88]. This could lead to accidents and serious injury. Also, there are so many attachments for many of the motorized wheelchairs that users are often overwhelmed and embarrassed by all the cadestry.

3.3.8 Costs.

The purchase cost of wheelchairs varies greatly depending on the type of wheelchair, and the accessories and customization needed. The standard wheelchair is the least expensive on the market and may cost between \$400 and \$1700. Sports chairs generally cost from \$800 for a racing model to \$1500 for a general sports model. The most expensive chair, the motorized chair, may cost from \$2000 to \$5000 [SHEP84], IROOP86].

3.4 Lack of Innovation.

Numerous factors inherent to the disabled population contribute to an absence of innovation in wheelchair design. Among them are an overall ignorance of the types of wheelchairs available, third-party reimbursement policies for the purchase of said vehicles, and the threat of product liability. An explanation of these factors is presented in this section.

3.4.1 Consumer Market.

In order to understand the elements which retard the innovation of wheelchair design, it is necessary to be cognizant of the consumer market and the manufacturing policies affecting the design. A 1977 survey revealed that one out of every two hundred Americans, approximately 1.2 million individuals, is a wheelchair user. Approximately half of these individuals reside in nursing homes, and this percentage is growing at a rate of 1.5 percent annually. In 1982, roughly 338,000 wheelchairs were sold in the United States for a total of 126 million dollars. It has been estimated that 250,000 of the wheelchairs sold annually are general-purpose, manual wheelchairs sold to rental agencies or institutions (e.g., nursing homes, or hospitals).

There are approximately fifty-three manufactures and importers of wheelchairs in the United States. However, Everest & Jeanings, Inc., Invacare Corporation and a few other large organizations dominate the market, constituting an oligopoly. The significance here is that there is limited competition stemming from the knowledge that an action by one manufacturer will constitute a retainatory action by the others.

3.4.2 Third-Party Reimbursers.

The wheelchair market is largely subjected to the doctrines of third-party reimbursers. Approximately balf of all wheelchairs purchased are at least partially funded by the government. Forty to forty-five percent are funded by private insurers, leaving only five to ten percent paid for in full by the user. The large percentage of externally funded purchases directs innovation to the types of devices which are eligible for reimbursement. Although the doctrines of insurers vary, all will pay for a wheelchair which is "medically necessary." Some agencies, such as the Veterans Administration and Medicaid, consider wheelchair accessories that psychologically benefit the user as medically necessary. However, many insurers will only pay for the least cost wheelchair which will meet the physical needs of the user. All insurers seem to employ the philosophy of price over performance in the funding of wheelchairs. In many cases, inexpensive wheelchairs will insur higher maintenance and repair costs than better quality, higher priced chairs. Unfortunately, this does not seem to be considered in the payment policies of third-party reimbursers. As a result, manufacturers are unlikely to be able to sell the higher priced models. This is a strong deterent to the manufacturing of such a chair.

3.4.3 Standards.

The FDA currently bas no official standards for wheelchairs. Performance standards are presently being developed by a delegation from the Rehabilitative Engineering Society of North America; however, compliance will be strictly voluntary.

The Veterans Administration, VA, does however, have standards for procured wheelchairs. Because the VA is the largest single purchaser of wheelchairs in the United States, manufacturers are obliged to conform to established guidelines. Unfortunately, however, until recently VA standards were developed with a particular wheelchair in mind. Actual performance standards were not adopted until 1977 for manual wheelchairs and 1982 for motorized wheelchairs. Until these standards were set, the VA guidelines essentially catered to the interest of the manufacturers. When standards were written in accordance with a particular model, it was in the manufacturer's best interest to produce chairs which conformed to the specifications of that model. This obviously curtailed innovation. In addition to this, a recent survey indicated that many manufacturers were unoure of the methodologies utilized by the VA in evaluating new products. This lack of knowledge makes innovation hazardous due to the uncertainty of whether products will meet VA standards, and if not, whether those standards minist be chanced.

3.4.4 Consumer Unawareness.

The selection of the correct wheelshair to purchase in usually the task of a desler or therapist. A physician's prescription in ordinarily required for third-party reimbursements, but physicians are generally not apprised to the various wheelshairs and accessories available. If the prescription does not indicate a particular brand, it is ordinarily left to the therapist or dealer to select a model. Most dealers, however, only handle the larger brands of wheelshairs. The decision concerning which brands to carry is usually dependent on quality, previous service, and profit margin. Dealers often buy large quantities of wheelshairs in order to receive substantial discounts and will generally only stock the models for which they were given the best price. Compounding this, they tend to carry wheelshairs with a higher percentage of profit. The lower priced manual wheelshairs carry a forty percent markup, while motorized wheelshairs generally have a thirty percent markup. Since most wheelshairs users are unaware of the various chairs and accessories available to them, they generally rely on the dealer's or therapist's recommendation and rarely order a special wheelshair.

3.4.5 Product Liability.

The threat of a product liability suit is another major obstacle to insovation.

According to one manufacturer, product liability suits have succeeded medical malpractice suits as the most opportunistic lawsuits filed today. The imminence of a lawsuit discourages physicians, dealers and therapists from recommending new products. This is especially true for entirely new products and less true for variations of existing products [SHEP84].

CHAPTER 4

DESIGN CRITERIA FOR AN AUTONOMOUS WHEELCHAIR

4.1 Objectives.

This design is expected to incorporate the optimal features of existing technology with certain innovative concepts to resolve problems associated with current motorized wheelchairs and to relieve the user from unnecessary tasks. It is the primary objective of this design to offer the disabled user a safe, reliable means of transportation. Secondary to this, but still of prime consideration, are the issues of assethetic value, reasonable costs and ease of use. Included are suggested modifications to the basic wheelchair design which should promote the safety and health of the user. Also proposed is the incorporation of a microprocessor-based control unit. The utilization of such a device would alleviate many navigational difficulties and allow simple customization for a user's residual capabilities through software. The criteria presented in this document are intentionally flexible so that the design engineers may incorporate the latest technological advances and modifications to further enhance the design.

It is the hope of this author that the criteria proposed here, along with each step of the design, he subjected to the review of potential users, rehabilitation engineers, and orthopedic surgeons as well as the various technical experts needed to implement the design. There would be no satisfaction in wasting valuable time and resources on research and development to create an aid which is either unwanted or impossible to finance.

4.2 Microprocessor Control Unit.

In order to achieve maximum capability, the proposed wheelchair should be microprocessor controlled. The utilization of a microprocessor control unit would eliminate many of the problems associated with present day motorized wheelchairs. It would also provide a significant number of additional facilities to the wheelchair user. Potential capabilities of this type of control are discussed in this section.

4.2.1 User Customization.

The hardware modifications needed to customize a motorized wheelchair for a user's residual capabilities are major contributors to the inordinate expense of the chair. For instance, a paraplegic generally has the ability to respond much more quickly in a given situation than the quadriplegic. This would indicate that the paraplegic would be capable of driving the wheelchair at greater speeds. For many motorized wheelchairs, this requires a hardware modification, a costly, time-consuming procedure. The wheelchair proposed in this document will utilize the microprocessor to allow the user or therapist to enter such customization variables. Likely parameters would include forward velocity, reverse velocity and maximum accelerations. These variables could easily be modified as the user's capabilities change.

4.2.2 Signal Processing.

Many disabled persons suffer from spastisticy or other conditions which cause unintentional or irregular motions. Such a motion involving the input control mechanism could be a source of inappropriate movement by a motorized wheelchair causing sudden acceleration or directional change. Such a movement could easily force the wheelchair to tin over. Through the use of a microprocessor, the user input signal could be processed to filter out erratic or inappropriate signals, resulting in a smoother, asfer ride. Digital filtering could be implemented for all users, regardless of their capabilities, to guard against moments of inattention or recklessness. However, the reduced response time may anany users who are not in need of the capability. Therefore, it is recommended that this feature be included as a customization option.

4.2.3 Closed Loop Control.

Many motorized wheelchairs use only the displacement of the user interface control mechanism (e.g. joystick) to determine velocity. This is referred to as open loop control hecause no feedhack data is collected or processed by the system. Closed loop (self-correcting) velocity control would be of great henefit to the wheelchair user. This would, in effect, he a cruise control facility for the wheelchair. The implementation of this control methodology would relieve the user of having to compensate for the incline when ascending or descending ramps. This could also prevent sudden acceleration or deceleration when driving from one type of surface to another. For instance, a wheelchair using open loop control could experience a sudden acceleration when passing from deep-pile carpet to a hardwood floor. Closed loop control provides a safety advantage, as well as convenience, and should he included in the design specifications for the proposed wheelchair.

4.2.4 Automatic Sensors

The use of automatic sensing devices has heen incorporated in the design of a motorized wheelchair at the Palo Alto Veterans Administration Medical Center. The sensing devices provide safety measures as well as navigational aids which are well worth considering in the design of the proposed wheelchair. The Veterans Administration "smart" wheelchair bas heen designed to incorporate ultrasonic sensors for safety and convenience purposes. Seasors mounted on the front of the chair detect obstructions in the user's path. If the obstacle is within a preprogrammed distance, the chair will slow down and stop to avoid a collision. The technology is responsive enough to sense the sudden opening of a door. An additional feature allows the chair to detect an object (person) moving slowly in front of it and slow down to follow that object.

Additional seasors mounted on each side of the chair allow the user to travel parallel to a wall with no user guidance. The "follow-that-wall" mode will detect and disregard doorways; however, if an opening of more that a few feet is detected, control of the wheelshair will automatically be returned to the user. The disregarding of small acertares allows the user to "follow" a nicket fence [JAFFR1].

In order to avoid reinventing the wheel, the technology utilized in the creation of the Palo Alto "smart" wheelchair should be investigated. An additional, similar feature which should be considered for the proposed wheelchair is the "follow-that-promo concept. A mode in which the wheelchair would travel parallel to a moving object could be extremely useful in allowing a wheelchair user to keep pace with a walking companion.

4.2.5 Path Memory.

A wheelchair designed at the Atlanta Veterans Administration Medical Center incorporates a path memory feature which allows the wheelchair to back out of confined spaces in which it is too difficult to maneuver. The feature utilizes closed loop control to measure and record distance and direction. If the user should encounter such an impediment, he would simply use the path memory feature to hack out of the confined space along the same route upon which he entered [KELL84]

The path memory feature should be incorporated into the design of the proposed wheelchair. It could be of great henefit to the handicapped individual, particularly to the severely disabled person who has little freedom of movement. It is suggested that the path memory feature he extended in the proposed wheelchair to, in effect, memorize routes which are frequently travelled by the user. This capability would relieve the user from unnecessary navigation when travelling familiar routes. A learning mode would be utilized when traversing a path for the first time in order to measure and record distances travelled in various directions. Upon arrival at the destination, the route would be associated with a unique key and stored for future reference. Once a route has been learned, it could be utilized in a playback mode to enable the user to once again travel that route. The chair would be automatically propelled to its destination according to the directions and distances previously recorded. If an unexpected, stationery obstacle was to be encountered, the chair would halt, and control would be returned to the user. If a moving obstacle was detected, the chair would slow down and follow that obstacle until it no longer presented a threat of collision. The playback mode could be used to either assist the user in travelling to a particular destination or retracing his path to the original starting position. A pause feature could be incorporated to allow for necessary moments of interrupted travel, such as the time spent in an elevator.

4.2.6 User Interface.

Severely disabled individuals can be stricken by a wide variety of debilitating conditions. Their residual capabilities vary so greatly that the selection of a single type of user interface to the proposed wheelchair would be unwise. It is recommended, therefore, that the microprocessor control unit be capable of accepting various types of input. The input could then be translated to a standard format for subsequent processins.

It should be noted that the standard wheelchair interfaces (i.e. joystick, sip-and-puff mechanism, etc.) limit the number of functions available to the user. It may be necessary, therefore, to incorporate the use of a scanning display in order to access the special capabilities of the chair. In this case, the user would continue to navigate bis chair through the standard interface and utilize a discrete switch to control the scanning display. The utilization of a less restrictive interface mechanism, such as a voice recognition unit, would eliminate the need for the display.

4.2.7 Personal Computer Interface.

An interface to a personal computer should be included in the design of the proposed chair. A physical link would provide the facility to transfer information between the one-board control unit and the personal computer. This would be useful in setting customization variables when the chair is newly purchased or when the capabilities of the user change. This could also provide extra storage for frequently travelled routes if the path memory feature is utilized. The physical link would be of great benefit, however, many disabled persons would be unable to establish that link without belp. Therefore, it is recommended that a remote link also be incorporated in the design. An infrared link could easily be established by the user by depressing a discrete switch mounted on the chair. Once control is transferred to the personal computer, a seanning or menu display could be incorporated to give the user full access to the computer's capabilities. This should include access to an environmental control system.

4.2.8 Miscellaneous.

Many severely disabled individuals are highly assecptible to pressure sores because they are unable to feel discomfort in their lower backs and legs. After being seated in one position for a lengthy period, healthy individuals feel discomfort and shift positions to compensate. Because many disabled persons can not feel this pain, they are unlikely to alter their positions and consequently develop pressure sores. To remedy this situation, the control unit of the proposed wheelebair could generate a tone at timed intervals to audibly remind the user to shift his weight.

Additional features which should be considered in the chair's design are as follows:

- an emergency alarm system to allow the occupant to call for help in the event of a mechanical failure,
- a failsafe switch to stop the chair and to temporarily shut down all motive properties of the chair in case of an emergency,
- · a battery voltage monitor to alert the user to low battery power, and
- a self-diagnostic system to detect possible failures in the components of the chair and alert the user to those possibilities.

4.3 Hardware.

The hardware stilized is the formation of the proposed wheelchair should incorporate the most reliable, economical and aesthetically pleasing components available at the time of construction. Consideration must be given to the fact that many severely disabled individuals essentially live in their chairs. They are largely dependent upon the reliable service of these chairs and would be greatly inconvenienced if a breakdown should occur. The appearance of their vehicles is also of great importance. Many motorized wheelchairs are assistentically offensive due to the massive number of attachments mounted on the chair. Many users spend the majority of their day in wheelchairs and could be embarrassed or even psychologically damaged if confined to an unsightly chair.

This section presents suggestions concerning the hardware components and mechanical design of the proposed wheelehair. This author does not presume to be mechanical expert nor do the criteria presented here cover the full range of hardware specifications required in a wheelehair design. These are merely ideas to be considered hased upon the extensive readings of the author.

4.3.1 Basic Design of Chair.

The mechanical design of the proposed wheelshair should incorporate modifications to the design of the standard chair. Except for the sports chair, the design of present day wheelshair is hasically the same as that of wheelshairs built fifty years ago. It is recommended that this design be challenged.

A likely candidate for examination would be the structure of the chair. The hasic structure of the wheelchair was not designed for the greater speeds (i.e. five to seven MPH) capable of today's high performance wheelchairs. As a result, chairs are more likely to tip over due to excessive sceeleration or deceleration, during high speed turns, or when ascending or descending inclines. Several of these problems may be solved by performance limitations programmed into the software which controls the motive force of the chair (e.g. maximum acceleration). However, preprogrammed performance limitations are unlikely to remedy all these situations. The obvious mechanical solution to the problem would he to lower the center of gravity and extend the wheel hase of the chair when travelling at higher speeds. Also to be considered in the design is a shift of the center of gravity when ascending or descending ramps. A forwards shift of the center of gravity would increase the stability against rearward tipover when ascending ramps. A rearwards shift would stabilize the chair against forward tipover when descending ramps. These shifts of the center of gravity could be automatically controlled by the on-board microprocessor.

The materials utilized in the construction of the seat and back of the wheelehair should sho be given prime consideration. Many present day wheelehairs have slung seats and backs constructed from fabric or plastic. This type of seating can promote pressure sores and postural deformities if used for extended periods of time. Because the proposed wheelehair is intended for severely disabled persons who are largely confined to their chairs, it is not advisable to use this type of seating. A solid seat and back would be much more suitable for preserving the health and well-being of the user. The factors listed below have been determined to promote good seating quality and should be considered in the design of the proposed wheelebair. The wheelehair seat unit should:

- · be adapted to the physique of the occupant,
- $\bullet\,$ stabilize the occupant's sitting posture,
- · enable the occupant to shift his weight,
- support the occupant's body weight and evenly distribute the pressure,
- maintain a comfortable temperature at the seat surface, and
- conduct moisture away from the surface of the seat [BRAT83].

4.3.2 Efficiency.

The design of the proposed wheelshair should be streamlined to obtain maximum efficiency in use. This could be largely accomplished through the examination of each component of the chair. Each unit which translates electrical battery power into locomotive energy should be examined for its efficiency, consistent with the objectives of reasonable costs, reliability and strength. Even components which do not directly affect the electrical system should be scrutinized. For example, studies have shown that smaller wheels are more efficient in use than larger ones. Since the wheelchair will be electrically driven, it is unaccessary to utilize the standard twenty-four inch wheels. Self-propulsion is the only purpose for these wheels, and motorized chairs are much too heavy to be procelled by the disabled user.

4.4 Modularity.

Extremely important in the consideration of design specifications should be the concept of modularity. The various features incorporated in this design should be separate and distinct so that the user is able to purchase exactly the options that he desires or that his residual capabilities dictate. This also implies that the microprocessor-based control functions associated with the proposed wheelchair be adaptable to existing motorized wheelchairs and their input interfaces (e.g. sip-and-puff mechanism, joystick, etc.). This would obviate the need for the user to purchase another wheelchair if he already owns one, and it would allow him to continue using the input interface which is best suited for his capabilities or with which he is most confectable.

4.5 Extensions.

Suggested extensions to the design of the proposed wheelchair are presented in this section. Although these ideas are considered reasonable and heneficial, the stiffing deterrents to innovation in wheelchair design dictate that these features not be included in the preliminary design. Before incorporation of these features is seriously contemplated, it should be ensured that a potential market actually exists and is able to finance the coions.

The incorporation of a "map" feature could be very hendicial for wheelchairs used in an institutional setting. The map would consist of a physical diagram of the building and would be stored in memory. The wheelchair user would simply indicate his destination and would he driven there automatically. The "map" concept was derived from a recent innovation in the automobile industry. This innovation features a visual monitor, connected to the automobile's dashboard, which displays a map of the city. When the driver indicates his destination, a potential route is displayed on the monitor. The current because of the automobile is determined through the use of a satellite and special sensing devices. Utilizing this equipment, an on-board microprocessor determines the location of the automobile and displays that location on the monitor. The display is refreshed as the automobile proceeds to its destination.

A map feature for the proposed wheelchair could be implemented using the same haise design. A map of the huilding stored internally would consist of the various distances and directions involved in travelling from one place to another. The current location of the wheelchair could be determined through sensing devices placed intermittently in the halls. For example, radio transmitters broadcasting unique frequencies might he placed in the corridors, allowing the chair to determine its location. From this information, the map could be used in determining the proper route to the destination. Personalized information concerning each wheelchair user could be stored on cassette tape to allow the wheelchair to be used by more than one person. Vital information might include the individual's room location, doctor's office, and gender (to determine which restroom was appropriate).

4.8 Conclusions.

This chapter has presented preliminary design criteria for an autonomous wheelchair. Navigational and safety problems associated with wheelchair usage have been addressed and potential solutions have been offered. The criteria presented in this document have been purposely open-ended to allow for further technology advances which may be incorporated in the final design. It can not be overly stressed that rehabilitation experts as well as the potential users themselves be consulted during each step of the design. Surveys should also be conducted to determine the potential market for the wheelchair and whether sufficient financing can be obtained for its purchase. Finally, it is highly recommended that the design specifications comply with the Veterans Administration's standards for procured wheelchairs and the FDA standards, once they are adopted.

CHAPTER 5

CONCLUSION

5.1 Deficiencies.

In spite of the recent progress in technological aids for the handicapped, there are many problems yet to be solved. Among them are the relatively high cost of equipment which is intended for a society largely unemployed, a general lack of knowledge concerning the availability and dissemination of said products, and the absence of human entirecting in the desira.

5.1.1 Acquisition.

The acquisition of technical aids by the duabled can be beset with many problems.

Although the price of computing facilities has dropped substantially in recent years,
the cost of adaptive equipment for the handicapped can still be quite high due to the
small consumer market. Many disabled persons are either unemployed or have low
paying jobs and can not afford expensive equipment. In many cases, third party
reimburners, such as Medicare or Medicaid, are charged with the responsibility of
financially adding the purchase of such equipment, but this can turn into a nightmar
or red tape. Several third party reimburners will only pay for the least cost aid which
is medically necessary and will not accept copayments by the disabled individual
[SHEP84]. Often, even finding the correct agency to contact can be difficult. The
Office of Technology Assessment has published a case study on health technology.

The following paragraph, cited from this study, refers specifically to the obstacles
faced by speech impaired individuals in obtaining financial assistance, but could easily
apply to most disabled persons.

There is an old axiom in medicine that when there are many different treatments for the same disorder the liciblicod is that none of them works very well. From the perspective of the severely physically diabbled nonzouch person, the same principle applies in finding a payment mechanism for the assistive communication device that will meet his needs: The many potential sources for lunding disguise the reality that reinhursement can be very difficult and sometimes impossible to obtain. Because no single agency in government or the impossible to obtain. Because no single agency in government or the tend to say it is not their responsibility and try to shift that

Also contributing to the expense of computer aids, is the high cost of research and development. Frequently these projects are funded by the government or agencies from the private sector. However, many devices which have been separately under still sell at the same approximate cost as similar devices offered by different manufacturers with no cost savings passed on to the consumer [PFROS5].

One of the most promising outlooks for the development of aids appears to be the practice of technology transfer. As with all types of information, technology can be passed from one type of application to another. In the case of the handicapped, the utilization of existing technology may belp to reduce the cost of research. For instance, a fully implantable, programmable insulin pump for diabetics was recently developed using NASA's technology for the Mars Viking space probe to release culture medium onto that plane's soil [EDMUST].

If technology developed for the disabled could also be utilized by the general population, the cost of resulting aids could be reduced significantly. For example, a tracking device has been developed which could assist the families of victims of Altheimer's disease to keep informed of their loved one's whereabouts. The device is worn on the wrist, much like an ordinary wristwatch, and generates signals which can be detected by special tracking equipment. If an Altheimer's sufferer should wander from home, he could be located through the use of this device. Utilization of the aid

could relieve the minds of many families, but the equipment is much too expensive to be put into limited use. Sadly, it will not be until the general population begins to use this device for keeping track of their pets that the price will be reduced sufficiently to be of use to the disabled.

5.1.2 Human Engineering.

With all the ferror over the benefits of computer aids, it is important to remain realistic about their advantages. Aids can only do just that - aid the user in overcoming his handicaps. There is not a device in existence today which will completely restore a function lost due to a severe impairment, and there is not likely to be in the near future. Utilization of a speech synthesizer does not mean that a person can talk and it does not serve anyone's best interest to pretend otherwise IBRADS41.

Most important in the design of technological sids in the buman engineering factor. So often, in large conferences and conventions, handicapped people are strictly for appearance only. It is so easy to listen to them speak, admire their course and come away touched by their accomplishments. But very little enlightment and understanding has taken place [GALL80]. It is too easy for engineers to begin the design of an aid without fully realizing the actual need of the disabled. For example, no one would design an aid which utilized a warning light for the blind user. However, how many engineers, when designs a portable aid for the blind would stop to realize that the blind individual essentially has only one band? The implication here is that the blind person will almost always have one hand occupied by a long cane or a guide dog, so that any aid which must be carried in both hands is virtually useless to him (GLLASSO)].

Although there has seemingly been great progress made in the development of aids for the handicapped, relatively few disabled persons have profited by it. Consider the following:

- Fewer than 0.3% of the more than 500,000 people with speech impairments use any type of auxiliary aid for communication,
- Fewer than one hundred automatic page readers have been placed in libraries for use by the blind, and
- Fewer than two thousand disabled persons customarily use telecommunications networks for the purpose of message or information exchange [TRIMS5].

These figures indicate that a very small percentage of the handicapped population is actually benefiting from computer technology. It is important to understand the causes of this under-utilization in order to resolve the associated problems and enable the handicapped to enjoy the services available to them. Several of the more apparent reasons include problems inherent to the handicapped population such as low income, third party reimbursement and small consumer market. But it is commonly felt that the problems go much deeper than this. Many handicapped individuals seem skeptical of the computer's usefulness in everyday life. If this is the case, then it is evident that the systems available today have not been designed to satisfy the demands of the disabled population. This implies that rehabilitation engineers must place more significance on understanding the needs of the disabled and applying the technology to appropriately solve them [TRIMS5]. For as E. L. Ghser has pointed out "Of what benefit is all that power, all that capability, when the one who needs it the most is denied it through a boughtless or incompetent design? [GLAS80)."

5.2 Summary.

The purpose of this document has been to inform the reader of the great progress which has been made in the area of computer asids for the handicapped. The advent of the microprocessor has reduced the size and expense of computing facilities effectuating a rice in the development of technological aids. Many disabled individuals may witness a reduction of handicap and consequently increased independence due to the availability of technical aids. Devices are currently in development to not only reduce the handicap of disabled persons but also to restore the impaired function using applied neural control.

This document has also presented a study of wheelchairs. A description of the evolution of wheelchairs was included along with a discussion of the problems resulting in the lack of innovation in wheelchair design. The purpose of this study was to establish the background for the design of an autonomous wheelchair.

In spite of the recent progress in advanced technological ads, there is still much work left to be done. By encouraging more disabled persons to directly contribute to the design and development of aids, problems with human engineering will be reduced significantly. The practice of technology transfer must be exploited to the full extent, and aids should be designed to incorporate further advancements in technology. This is essential for continuing the progress in computer aids for the handicapped. For as C. Emerson has reminded us, "... we should remember that the disabled are the only minority group anyone can join BURINSS."

- BOZZ81 Bozzuto, R. C. "The Universal Translating Modem: An Advanced Telecommunication Device For The Deaf" Proceedings of the Johns Hopkins First National Search For Applications of Personal Computing To Aid The Handicapped, pages 62-64. October 31, 1981.
- BRAD85 Brady, M. E. "The State of The Art of High Tech Aids" In Technology For Disabled Persons: Conference Papers, Discovery 84. pages 155-164. University of Wisconsin-Stout, Menomonie, Wisconsin, 1985.
- BRAT83 Braattgard, S. O., Lindstrom, I., Severinsson, K., and Wihk, L. "Wheelchair Design and Quality" Scandinavian Journal of Rehabilitation Medicine, Suppl. 9, pages 15-19, 1983.
- BURN85 Burnett, N., and Neimark, J. "DP and The Disabled; The Computers Currently On The Market Are Essentially Hostile To The Handicapped" Datamation, pages 22-30, January 1, 1985.
- BRUBS6 Bruhaker, C. E., Ph.D. "Wheelchair Prescription: An Analysis of Factors That Mobility And Performance" Journal Of Rehabilitation Research And Development, Vol. 23, No. 4, pages 19-29. October 1986.
- CHIZ85 Chizeck, H. J. "Helping Paraplegies Walk: Looking Beyond The Media Blitz" Technology Review, pages 55-63. July, 1985.
- COLL84 Collins, C. C., and Deering, M. F. "A Microcomputer Based Blind Mohility Aid" In Frontiers of Engineering And Computing In Health Care - 1984 Proceedings - Sixth Annual Conference IEEE Engineering In Medicine and Biology Society, pages 52-55, IEEE, New York, NY, 1984.
- DONA83 Donaldson, P. E. K. "Engineering Visual Prostheses" Engineering in Medicine and Biology, Vol. 2, No. 2, pages 14-18, June 1983.
- DUFF81 Duffy, J. F. "A Video Voice Display For Use In Speech Training For The Deal" Proceedings of the Johns Hopkins First National Search For Applications of Personal Computing To Aid The Handicapped, pages 23-25. October 31, 1981.
- DURI78 Durie, N. D. "A Single-Switch Control for Wheelchairs And Other Equipment" Medical Progress Through Technology, Vol. 6 pages 15-18. 1978.
- DURR84 Durre, K. P., Friehoff, T. M., Schmidt-Lademann, F. "Braillehutler: A Successful Micro-Computer Based Aid For Mainstreaming Blind Children" Proceedings of the Third Annual Workshop On Computers And The Handicapped, pages 17-22. IEEE Computer Society Press, Silver Spring, MD 1984.
- EDMU87 Edmunds, L. "Space Age Insulin Pump Implanted in Human" Johns Hopkins Magazine, Vol. XXXIX, No. 1, page 46, February 1987.
- EDWA87 Edwards, D. D. "Good Connections? It's In The Chips" Science News, Vol. 131, No. 6, page 86. February 7, 1987.

- FREE84 Freeston, I. L., Callaghan, V. L., and Russell, N. D. "A Portable Navigation Aid For The Bind" In Frontiers of Engineering And Computing In Health Care - 1924 Proceedings - Sixth Annual Conference IEEE Engineering In Medicine and Biology Society, pages 247-249. IEEE, New York, NY, 1982.
- GALL80 Gallion, D. R. "The Johnstown Experience: An Adventure in Awareness Training" Proceedings of the IEEE Computer Society Workshop On The Applications Of Personal Computing To Aid The Handicapped, pages 58-60. IEEE Publishing Services, New York, NY 1980.
- GLAS80 Glaser, E. L. "Human Engineering For The Physically Handicapped" Proceedings of the IEEE Computer Society Workshop On The Applications Of Personal Computing To Aid The Handicapped, pages 11-13. IEEE Publishing Services, New York, NY 1980.
- GOLD81 Goldenberg, E. P. "Flexible High Bandwidth Communication For Motorically Impaired Persons" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicanned, pages 190-192. October 31, 1981.
- GRAUS Graupe, D., Ph.D., Kohn, K. H., M.D., Bassens, S., M.S., and Naccarato, E. M.S. "Controlling Electrical Stimulation Of Paraphegics Via EMG-Based Posture-Mapping," In Frontiers of Engineering And Computing In Health Care 1984 Proceedings Sixth Annual Conference IEEE Engineering In Medicine and Biology Society, pages 4-8. IEEE, New York, NY 1984.
- HAGG83 Haggard, M. P. "Introduction" High Technology Aids For The Disabled, page 69. Butterworth & Co. Ltd. 1983.
- HATO81 Haton, M. C., and Haton, J. P. "Computer Aided Speech Analysis And Training For The Hearing Impaired" Proceedings of the IFIP-IMIA Working Conference On Uses of Computers In Aiding The Disabled, pages 188-198. North Holland Publishing Company New York, NY 1981.
- HEYES3a Heyes, A. D., Dodds, A. G., Carter, D. D. C., and Howarth, C. I. "Evaluation of the Mobility of Blind Pedestrians" High Technology Aids For The Disabled, pages 14-19, Butterworth & Co. Ltd. 1983.
- HEYES3b Heyes, A. D. "The Sonic Pathfinder A New Travel Aid For The Blind" High Technology Aids For The Disabled, pages 105-171. Butterworth & Co. Ltd. 1983.
 - HITC83 Hitchcock, E. R. "Neural-Implantation Techniques For Disability" High Technology Aids For The Disabled, pages 71-76. Butterworth & Co. Ltd. 1983.
 - HOGA83 Hogan, B. J. "Voice Operated Controller Helps Quadraplegic Operate Appliances" Design News, pages 70-71, August 8, 1983.
 - HOLLS1 Holladay, D. "Braille-Edit Program: Connecting An Apple II Computer With A Versabraille Paperless Brailler" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapoed nages 231-233. October 31, 1981.

- HOOP86 Hooper, E. "Have I Got A Chair For You" Disabled USA, Vol. 1, No. 2, page 26-29, 1986.
- HOLMS1 Holman III, F. S. "Communications, Environment Controller, and Music Synthesizer/Color Graphics Generator" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 138-139. October 31, 1981.
- HOROSI Horowitz, B. R., Streepey, B. M., and Ton, D. T. "PALTALK: A Spelled-Speech Talking Interface For A Computer Terminal" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 261-266. October 31, 1981.
- JAFF81 Jaffe, D. L. "An Ultrasonic Head Position Interface For Wheelchair Control" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 142-145. October 31, 1981.
- JOHN81 Johnson, A.B., and Hagstad, R. F. "DTMF Telecommunications For The Deaf And Speech Impaired" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 29-31, October 31, 1981.
- JONES1 Jones, R. L. "Row/Column Scanning With A Dynamic Matrix" In Proceedings of The Johns Hopkins Firs National Search For Applications of Personal Computing To Aid The Handicapped, pages 6-8. October 31, 1981. University of Wisconsin-Stout, Menomonic, Wisconsin, 1985.
- KAME86 Kamenetz, H. "Wheelchairs and Other Indoor Vehicles For The Disabled" Orthotics Etcetera 3rd edition, Baltimore, MD:Williams & Wilkins, pages 464-517, 1986.
- KANESI Kanevsky, D., and Skurkovich, G. "A Vibrotactual Articulator" Proceedings of the IFIP-IMIA Working Conference On Uses of Computers In Aiding The Disabled, pages 273-277. North Holland Publishing Company New York, NY 1981.
- KELL84 Kelly, G. W., Ackerman, T. M., and Moody, L. E. "Sonic Orientation And Navigational Aid (SONA)" In Frontiers of Engineering And Computing In Health Care - 1884, Proceedings - Sixth Annual Conference IEEE Engineering In Medicine and Biology Society, pages 236-238. IEEE, New York, NY, 1892.
- KIHN81 Kihneman, D. C., and Salathiel, R. M. "Computer Fingerspelling" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 33-34. October 31, 1981.
- KULI84 Kulikowski II, S. "Integrating Scanning Software: Training and Vocational Uses For The Nonvocal" Proceedings of the Third Annual Workshop On Computers And The Handicapped, pages 35-40. IEEE Computer Society Press, Silver Spring, MD 1984.

- KURZ81 Kurzweil, R. C. "Kurzweil Reading Machine For The Blind" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 236-237, October 31, 1981.
- LEIF81 Leifer, L. "Restoration of Motor Function A Robotics Approach" Proceedings of the IFIP-IMIA Working Conference On Uses of Computers In Helping The Disabled, pages 3-17. North Holland Publishing Company New York, NY 1981.
- MCKI84 McKibbin, W. L. "High Tech For the Handicapped: An Interview With Susan Phillips Development Director of Employment Programs, Sensory Aids Foundation" Infosystems, Vol. 31, page 52, January 84.
- MCWI84 McWillaims, P. A. Personal Computers And The Disabled, Quantum Press, Doubleday & Co., Garden City, NY, 1984
- MILLS5 Miller, G. E., Brown, T. E., and Randolph, W. R. "Voice Controller For Wheelchairs" Medical and Biological Engineering and Computing, Vol. 23, pages 597-600, 1985.
- MORTS3 Mortimer, J. T. "Electrical Excitability: The Basis For Applied Neural Control" Engineering in Medicine and Biology, Vol. 2, No. 2, pages 12-13.
- June 1983.

 MOYESI Moyer, J. J. "An overview of Some Applications of Sophisticated Electronics Technology To Visual Handicaps" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicamed, cases 28–30. October 31, 1987.
- NAKA84 Nakano, Y., Fujie, M. and Hosada, Y. "Hitatehi's Robot Hand" Robotic Age, Vol. 6, No. 7, pages 18-20. July 1984.
- OTA83 Health Technology Gase Study 86: Assistive Devices for Severe Speech Impairments, Washington, D.C.: U.S. Congress, Office of Technology Assessment, OTA-HGS-26 December 1983.
- PARK83 Parkins, C. W. "Cochlear Implant: A Sensory Prosthesis Frontier" Engineering in Medicine and Biology, Vol. 2, No. 2, pages 18-27. June 1983
- PFRO85 Pfrommer, M. C. "Utilization Of Technology: Consumer Perspective" In Technology For Disabled Persons: Conference Papers, Discovery 84, pages 237-242. University of Wisconsin-Stout. Menomonie. Wisconsin. 1985.
- ROSES1 Rosen, S., Fourcin, A. H., Walliker, J. R., Douek, E. E., and Moore, B. C. J. "External Electrical Stimulation of the Cochlea In The Totally Deal" Proceedings of the IFF-IMIA Working Conference On Uses of Computers In Aiding The Disabled, pages 107-183. North Holland Publishing Company New York, NY 1981.
- ROSE85 Rosen, M. J. "The Nan Davis Story: A Trail of False Hopes" Technology Review, pages 60-61, July 1985.

- ROSO81 Rosov, R. J., Parikh, Y. B., and Gavin, W. J. "A Wearable Tactual Vocoder For Profoundly Deaf Children" Proceedings of the 37th Annual Conference On Engineering In Medicine and Biology, page 307. Alliance for Engineering in Medicine and Biology, Publishers Bethesda, MD 1984.
- ROSS79 Rossier, A. B., Sarkarati, M., Crawford, G. E., and Dietz, J. "Development of a Wheelchair Using A Mycelectric Control System" Bulletin of Prosthetics Research, BPR 10-31, pages 113-116 Spring 1979.
 ROSS81 Ross, D. A. "A Musical Language Computer Terminal For The Visually
- Impaired" Proceedings of the Johns Hopkins First National Search For Applications of Personal Computing To Aid The Handicapped, pages 197-199. October 31, 1981.
- SCHE84 Schetky, L. McD. "Shape Memory Effect Alloys For Robotic Devices" Robotic Age, Vol. 6, No. 7, pages 13-17. July 1984.
- SCHN80 Schneider, W., Seamone, W., and Schmeisser, G. "A Microprocessor-Controlled Robotic Arm Allows Self-Feeding For A Quadraplegic" Proceedings of the IEEE Computer Society Workshop On The Applications Of Personal Computing To Aid The Handicapped, pages 31-36. IEEE Publishing Services, New York, NY 1980.
- SCHN81 Schneider, W., Schmeisser, G., and Seamone, W. "A Computer-Aided Robeic Arm/Worktable System for the High-twel Quadraplegie" Computer, Vol. 14, No. 1, pages 41-47. January 1981.
 - SCHO81 Schofield, J. M. Microcomputer-Based Aids For The Disabled, London, UK: Heyden and Son, Inc., 222 pages. 1981.
- SEAM79 Seamone, W., and Schmeisser G. "Interdisciplinary Development and Evaluation of Externally Powered Upper-Limb Prostbeses and Orthoses" Bulletin of Prosthetics Research, BPR 10-31, pages 78-85. Spring 1979.
- SEAM81 Seamone, W., and Schmeisser G. "New Control Techniques For Wheelchair Mobility" Johns Hopkins APL Technical Digest, Vol. 2, No. 3, pages 179-184. 1981.
- SEAM83 Seamone, W., and Schmeisser, G. "Wheelchair Control And Robotic Arm/Worktable System For High-Spinal-Cord-Injured Persons" Rehabilitation Research & Development Progress Reports, pages 61-62.
- SELD83a Edited by Seldon, P., and Todd, T. "Development And Evaluation Of A Robotic Aid For The Severely Disabled" Rehabilitation Research & Development Progress Reports, pages 62-63, 1983.
- SELD83b Edited by Seldon, P., and Todd, T. "Modifications To Wheelchair For A High-Level Quadraplegic" Rehabilitation Research & Development Progress Reports, page 133, 1983.
- SELD83c Edited by Seldon, P., and Todd, T. "A Wheelchair Bumper" Rehabilitation Research & Development Progress Reports, pages 133-134. 1983.

- ELD83d Edited by Seldon, P., and Todd, T. "Adjustable Weights To Prevent Wheelchair Tipping" Rehabilitation Research & Development Progress Reports, page 134, 1983.
- SHEN84 Shennib, A., and Kondraske, G. V. "An Improved Telecommunication Aid For The Deal" in Prontiers of Engineering And Computing in Health Care - 1983 Proceedings - Sixth Annual Conference IEEE Engineering In Medicine and Biology Society, pages 38-40. IEEE, New York, NY, 1984.
- SHEP84 Shep, D.S., Karon, S.L. Health Technology Case Study 30: The Market for Wheelchairs: Innovations and Federal Policy, Washington, D.C.: U.S. Congress, Office of Technology Assessment, OTA-HCS-30 November 1984.

Siebert, T. W. "Computer Aided Braille Trainer" In Frontiers of

Engineering And Computing In Health Care - 1984 Proceedings - Sixtb Annual Conference IEEE Engineering In Medicine and Biology Society, pages 243-246. IEEE, New York, NY. 1984.

SIBE84

- SLAT86 Slater, M. "Let Your Voice Do The Walking" Disabled USA, Vol. 1, No. 2, pages 19-21, 1986.
- THOM83 Thoma, H., Frey, M., Holle, J., Kern, H., Reiner, E., Schwanda, G., and respective for the property of the property
- TILL81 Till, J. A., and Maier, M. F. "Numbered-Accessed Expressive Language Program" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 27-28. October 31, 1981.
- TRIM85 Trimble, J., and Lambert, R. W. "Perspectives On Computers and Disable Persons" Proceedings of the 38th Annual Conference on Engineering in Medicine and Biology, page 311. The Alliance for Engineering in Medicine and Biology, Publisher Washington DC. 1985.
- TYND87 Tyndall, K. "Flexible Robots" Insight, page 49. April 13, 1987.
- VILL81 Villehur, E. "Signal Processing For Hearing Aids" Proceedings of the IFIF-IMIA Working Conference On Uses of Computers in Aiding The Disabled, pages 217-229. North Holland Publishing Company New York, NY 1981.
- WALS81 Walsh III, W. M. "An Emergency Deaf Communications System" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 35-36. October 31, 1981.
- WEISS5 Weiss, L. H. "An Improved Row/Column Scanning System" In Technology For Disabled Persons: Conference Papers, Discovery 84, pages 88-92. University of Wisconsin-Stout, Menomonie, Wisconsin, 1985.

- WHIT81 White, R. L., Herndon, M. K., and Mathews, R. G. "The Use of Computers in The Development of a Cochiever Prosthesis" Proceedings of the FIP-IMIA Working Conference On Uses of Computers In Aiding The Disabled, pages 147-165. North Holland Publishing Company New York, NY 1981.
- ZING81 Zingis, A. "Paperless Braille Micro" Proceedings of the Johns Hopkins First National Search For Applications Of Personal Computing To Aid The Handicapped, pages 267-226. October 31, 1981.

APPENDIX

GLOSSARY OF TERMS

ALS Amyotrophic Lateral Sclerosis, A syndrome marked by

muscular weakness and atrophy with spasticity due to degeneration of motor neurons of spinal cord, medulla, and cortex. Commonly known as Lou Gehrig's disease.

Alzheimer's disease Presenile dementia, which is similar to senile dementia but

occurs in the 40 - 60 year age group. The disease has a relentless and irreversible course but may take from a few months to four or five years to go to the stage of complete helplessness.

athetoid Affected with athetosis.

athetosis A condition wherein there are slow, irregular, twisting, snakelike muscular movements seen mostly in the upper

extremities, especially in the bands and fingers.

atrophic Pertaining to or marked by atrophy.

atrophied Wasted, Afflicted with atrophy.

atrophy Decrease in size or wasting away of a body part or tissue.

braille A system of writing for the blind that uses characters made up of raised dots.

eachectic Affected by cachexia.

cachexia General physical wasting and malnutrition usually associated

closed loop An automatic control system for an operation or process in

which feedback in a closed path or group of paths acts to maintain output at a desired level.

cochlea A winding cone-shaped tube forming a portion of the inner ear.

It contains the organ of Corti, the receptor for hearing.

An outer layer of an organ or part, as of the kidney.

cortex An outer layer of an organ or part, as of t cutaneous Of, pertaining to, or affecting the skin.

DTMF signals Dual Tone Multiple Frequency signals - Touch Tone signals.

dearth An inadequate supply.

disability

Any restriction or lack of ability (resulting from an impairment) in performing an activity in the manner or within the range considered normal for a human being.

electrode A collector or emitter of electric charge or electric-charge

carriers, as in a semiconducting device.

fasciculation Muscular twitching involving contiguous groups of muscle

Hiboro.

FNS Functional Neuromuscular Stimulation - The practice of

electrically exciting muscles to restore some use to limbs

paralyzed due to nerve damage.

gout Metabolic disease marked by acute arthritis and inflammation of the joints. Joints affected may be at any location but gout usually begins in the knee or foot.

handicap A disadvantage for a given individual, resulting from an

impairment or disability, that limits or prevents the fulfillment of a role that is normal for that individual.

role that is normal for that individual.

impairment Any loss or abnormality of psychological, physiological or

anatomical structure or function.

The underlying foundation or basic framework.

lesion An abnormal change in structure of an organ or part due to

injury or disease.

Pertaining to electrical properties of the muscles.

neural Of relating to, or affecting a nerve or the nervous system.

neuromuscular Concerning both nerves and muscles.

orthosis The straightening or correction of a deformity or disability.

open loop A control system for an operation or process in which there is no self-correcting action as there is in a closed loop.

paraplegia Paralysis of lower portion of the body and of both legs.

An instrument that records the distance a walker covers by

responding to bis body motion at each step.

percutaneous Passed, done, or effected through or by means of the skin.

Phosphene A luminous impression due to excitation of the retina or visual

cortex.

pressure sore

A sore cause by pressure from a splint or other appliance, or from the body itself when it bas remained immobile for extended periods of time.

prosthesis Replacement of a missing part by an artificial substitute.

quadriplegia Paralysis affecting all four limbs.

scoliosis Lateral curvature of the spine. Usually consists of two curves, the original one and a compensatory curve in the opposite direction.

spasticity Hypertension of muscles causing stiff and awkward movements: the result of upper motor neuron lesion. transcutaneous Passing or entering through the skin (as in an inoculation).

transducer A device that is actuated by power from one system and supplies power usually in another form to a second system.
tactile Of or relating to the sense of touch.

tactile Of or relating to the sense of touch

ADVANCED TECHNOLOGICAL AIDS FOR THE HANDICAPPED

bv

Leslie Howe Marley

B.S., Virginia Polytechnic Institute and State University, 1979

AN ABSTRACT OF A MASTER'S REPORT

submitted in partial fulfillment of the

requirements for the degree

MASTER OF SCIENCE

Department of Computing and Information Sciences

KANSAS STATE UNIVERSITY Manhattan, Kansas There is a significant number of bandicapped individuals in society today, and their number is ever increasing. Included in this number are those who suffer from motor disabilities, blindness, deafness, and speech impairment. These disabilities may be caused by congenital defects, accidents, war, degenerative nerve and muscle diseases and numerous other reasons. The computer has long been considered a potential aid for the bandicapped, but it was not until the advent of the microprocessor that it has become inexpensive and compact enough to be of any practical value.

This paper establishes the need for advanced technological aids for the handicapped. An overview of the state-of-the-art in these sides is presented, making a distinction between aids which reduce the handicap and those which replace the impaired function using technological methods. The paper also presents a study of wheelechairs, both manual and automated, and concludes with the design criteria for an autonomous wheelechair. It is proposed in this design to utilize available technologies which represent low cost, reliable solutions in order to provide the most sophistication and capability at the minimum cost.